

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025894.D
 Acq On : 17 Dec 2021 18:38
 Operator : JC/MD
 Sample : PB141495ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141495ZHE#02

Quant Time: Dec 19 02:34:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219135	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89267	50.613	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.220%		
35) Dibromofluoromethane	5.391	113	70684	46.251	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.500%		
50) Toluene-d8	8.653	98	273285	49.079	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.160%		
62) 4-Bromofluorobenzene	11.085	95	95163	44.990	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 89.980%		
Target Compounds						
20) Methylene Chloride	2.794	84	6485	1.975	ug/l	97
43) Isopropyl Acetate	6.348	43	13223	3.564	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025894.D
Acq On : 17 Dec 2021 18:38
Operator : JC/MD
Sample : PB141495ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141495ZHE#02

Quant Time: Dec 19 02:34:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

