

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
 Data File : VX025897.D
 Acq On : 17 Dec 2021 19:49
 Operator : JC/MD
 Sample : PB141495TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141495TB

Quant Time: Dec 19 02:35:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228563	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100021	54.422	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.840%			
35) Dibromofluoromethane	5.391	113	78764	49.412	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.820%			
50) Toluene-d8	8.653	98	297353	51.198	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.400%			
62) 4-Bromofluorobenzene	11.085	95	104726	47.469	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.940%			
Target Compounds						
20) Methylene Chloride	2.788	84	6532	1.824	ug/l	97
25) 2-Butanone	4.574	43	2150	1.739	ug/l	92
43) Isopropyl Acetate	6.349	43	12715	3.286	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025897.D
Acq On : 17 Dec 2021 19:49
Operator : JC/MD
Sample : PB141495TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141495TB

Quant Time: Dec 19 02:35:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

