

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121821\
 Data File : VX025877.D
 Acq On : 17 Dec 2021 11:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 02:25:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	121	0.00
2 T	Dichlorodifluoromethane	50.000	55.687	-11.4	141	0.00
3 P	Chloromethane	50.000	46.918	6.2	111	0.00
4 C	Vinyl Chloride	50.000	46.711	6.6#	116	0.00
5 T	Bromomethane	50.000	44.244	11.5	108	0.00
6 T	Chloroethane	50.000	44.554	10.9	115	0.00
7 T	Trichlorofluoromethane	50.000	46.377	7.2	121	0.00
8 T	Diethyl Ether	50.000	47.998	4.0	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.806	2.4	128	0.00
10 T	Methyl Iodide	50.000	49.768	0.5	111	0.00
11 T	Tert butyl alcohol	250.000	223.204	10.7	113	0.00
12 CM	1,1-Dichloroethene	50.000	45.677	8.6#	113	0.00
13 T	Acrolein	250.000	221.914	11.2	115	0.00
14 T	Allyl chloride	50.000	44.762	10.5	108	0.00
15 T	Acrylonitrile	250.000	245.765	1.7	114	0.00
16 T	Acetone	250.000	286.493	-14.6	121	0.00
17 T	Carbon Disulfide	50.000	38.998	22.0	98	0.00
18 T	Methyl Acetate	50.000	49.792	0.4	114	0.00
19 T	Methyl tert-butyl Ether	50.000	50.235	-0.5	114	0.00
20 T	Methylene Chloride	50.000	49.011	2.0	110	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.199	7.6	110	0.00
22 T	Diisopropyl ether	50.000	50.008	-0.0	115	0.00
23 T	Vinyl Acetate	250.000	260.412	-4.2	115	0.00
24 P	1,1-Dichloroethane	50.000	46.426	7.1	110	0.00
25 T	2-Butanone	250.000	256.976	-2.8	117	0.00
26 T	2,2-Dichloropropane	50.000	45.831	8.3	112	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.237	5.5	112	0.00
28 T	Bromochloromethane	50.000	42.666	14.7	94	0.00
29 T	Tetrahydrofuran	250.000	249.808	0.1	114	0.00
30 C	Chloroform	50.000	46.410	7.2#	110	0.00
31 T	Cyclohexane	50.000	50.113	-0.2	129	0.00
32 T	1,1,1-Trichloroethane	50.000	44.789	10.4	108	0.00
33 S	1,2-Dichloroethane-d4	50.000	44.242	11.5	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	113	0.00
35 S	Dibromofluoromethane	50.000	46.318	7.4	102	0.00
36 T	1,1-Dichloropropene	50.000	48.159	3.7	112	0.00
37 T	Ethyl Acetate	50.000	53.410	-6.8	116	0.00
38 T	Carbon Tetrachloride	50.000	44.604	10.8	106	0.00
39 T	Methylcyclohexane	50.000	53.591	-7.2	136	0.00
40 TM	Benzene	50.000	50.295	-0.6	111	0.00
41 T	Methacrylonitrile	50.000	52.999	-6.0	117	0.00
42 TM	1,2-Dichloroethane	50.000	50.204	-0.4	110	0.00
43 T	Isopropyl Acetate	50.000	51.919	-3.8	114	0.00
44 TM	Trichloroethene	50.000	49.569	0.9	113	0.00
45 C	1,2-Dichloropropane	50.000	51.648	-3.3#	115	0.00
46 T	Dibromomethane	50.000	49.801	0.4	111	0.00
47 T	Bromodichloromethane	50.000	47.146	5.7	103	0.00
48 T	Methyl methacrylate	50.000	54.364	-8.7	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121821\
 Data File : VX025877.D
 Acq On : 17 Dec 2021 11:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 19 02:25:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1056.780	-5.7	112	0.00
50 S	Toluene-d8	50.000	46.588	6.8	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	266.203	-6.5	116	0.00
52 CM	Toluene	50.000	51.231	-2.5#	111	0.00
53 T	t-1,3-Dichloropropene	50.000	50.741	-1.5	105	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.166	-2.3	109	0.00
55 T	1,1,2-Trichloroethane	50.000	51.361	-2.7	114	0.00
56 T	Ethyl methacrylate	50.000	55.170	-10.3	116	0.00
57 T	1,3-Dichloropropane	50.000	51.489	-3.0	113	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	256.703	-2.7	124	0.00
59 T	2-Hexanone	250.000	275.820	-10.3	118	0.00
60 T	Dibromochloromethane	50.000	46.848	6.3	99	0.00
61 T	1,2-Dibromoethane	50.000	51.048	-2.1	113	0.00
62 S	4-Bromofluorobenzene	50.000	47.869	4.3	109	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	112	0.00
64 T	Tetrachloroethene	50.000	49.230	1.5	112	0.00
65 PM	Chlorobenzene	50.000	51.882	-3.8	114	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.867	2.3	106	0.00
67 C	Ethyl Benzene	50.000	52.471	-4.9#	113	0.00
68 T	m/p-Xylenes	100.000	107.018	-7.0	112	0.00
69 T	o-Xylene	50.000	54.176	-8.4	114	0.00
70 T	Styrene	50.000	54.524	-9.0	111	0.00
71 P	Bromoform	50.000	47.194	5.6	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	113	0.00
73 T	Isopropylbenzene	50.000	50.646	-1.3	117	0.00
74 T	N-ethyl acetate	50.000	54.959	-9.9	120	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.336	-0.7	115	0.00
76 T	1,2,3-Trichloropropane	50.000	48.855	2.3	113	0.00
77 T	Bromobenzene	50.000	50.520	-1.0	114	0.00
78 T	n-propylbenzene	50.000	51.445	-2.9	118	0.00
79 T	2-Chlorotoluene	50.000	48.764	2.5	113	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.585	-3.2	115	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.909	12.2	97	0.00
82 T	4-Chlorotoluene	50.000	50.005	-0.0	113	0.00
83 T	tert-Butylbenzene	50.000	51.554	-3.1	120	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.102	-4.2	114	0.00
85 T	sec-Butylbenzene	50.000	54.039	-8.1	123	0.00
86 T	p-Isopropyltoluene	50.000	54.712	-9.4	123	0.00
87 T	1,3-Dichlorobenzene	50.000	51.313	-2.6	117	0.00
88 T	1,4-Dichlorobenzene	50.000	49.938	0.1	115	0.00
89 T	n-Butylbenzene	50.000	55.708	-11.4	129	0.00
90 T	Hexachloroethane	50.000	43.318	13.4	100	0.00
91 T	1,2-Dichlorobenzene	50.000	52.523	-5.0	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.199	7.6	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.731	-7.5	132	0.00
94 T	Hexachlorobutadiene	50.000	60.351	-20.7	147	0.00
95 T	Naphthalene	50.000	52.843	-5.7	121	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121821\
Data File : VX025877.D
Acq On : 17 Dec 2021 11:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 19 02:25:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.178	-8.4	129	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6