

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
 Data File : VX039307.D
 Acq On : 18 Dec 2023 13:32
 Operator : JC/MD
 Sample : 05806-10ME
 Misc : 5.23g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EOLROME

Quant Time: Dec 18 21:46:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 18 21:43:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	243955	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	215168	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	100764	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	49598	25.921	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	51.840%#
7) Chloroethane-d5	1.648	69	20783	14.695	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	29.380%#
11) 1,1-Dichloroethene-d2	2.276	65	25982	30.933	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.860%
21) 2-Butanone-d5	4.489	46	140616	81.185	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.180%
24) Chloroform-d	5.056	84	123974	36.600	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	73.200%
26) 1,2-Dichloroethane-d4	5.958	65	91035	41.499	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.000%
32) Benzene-d6	5.970	84	251922	37.937	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.880%
36) 1,2-Dichloropropane-d6	7.312	67	83244	38.953	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.900%
41) Toluene-d8	8.647	98	219216	36.428	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.860%#
43) trans-1,3-Dichloroprop...	8.951	79	35520	32.662	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.320%
47) 2-Hexanone-d5	9.397	63	99287	77.787	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.790%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	116764	38.516	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.040%
66) 1,2-Dichlorobenzene-d4	12.323	152	78585	41.260	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
Data File : VX039307.D
Acq On : 18 Dec 2023 13:32
Operator : JC/MD
Sample : 05806-10ME
Misc : 5.23g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EOLROME

Quant Time: Dec 18 21:46:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 18 21:43:41 2023
Response via : Initial Calibration

