

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
 Data File : VX039331.D
 Acq On : 18 Dec 2023 23:08
 Operator : JC/MD
 Sample : 05835-13 100X
 Misc : 2.05g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COAR0

Quant Time: Dec 19 04:21:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 18 22:55:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	207262	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	191985	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	90842	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	54678	33.635	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.260%
7) Chloroethane-d5	1.666	69	44725	37.221	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.440%
11) 1,1-Dichloroethene-d2	2.307	65	28459	39.881	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.760%
21) 2-Butanone-d5	4.446	46	156097	106.078	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.080%
24) Chloroform-d	5.056	84	130114	45.213	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.420%
26) 1,2-Dichloroethane-d4	5.958	65	93685	50.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.540%
32) Benzene-d6	5.977	84	259204	43.747	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.500%
36) 1,2-Dichloropropane-d6	7.306	67	88644	46.488	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.980%
41) Toluene-d8	8.647	98	224688	41.846	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.700%
43) trans-1,3-Dichloroprop...	8.952	79	38133	39.299	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.600%
47) 2-Hexanone-d5	9.385	63	111582	97.976	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.980%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	129512	47.879	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.760%
66) 1,2-Dichlorobenzene-d4	12.317	152	79004	46.011	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.020%
Target Compounds						
					Qvalue	
51) Chlorobenzene	10.080	112	8538	2.136	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
Data File : VX039331.D
Acq On : 18 Dec 2023 23:08
Operator : JC/MD
Sample : 05835-13 100X
Misc : 2.05g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AR0

Quant Time: Dec 19 04:21:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 18 22:55:37 2023
Response via : Initial Calibration

