

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
 Data File : VX039305.D
 Acq On : 18 Dec 2023 12:45
 Operator : JC/MD
 Sample : 05806-03ME
 Misc : 6.21g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E0LR6ME

Quant Time: Dec 18 21:45:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 18 21:43:41 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 12/19/2023
 Supervised By : Mahesh Dadoda 12/19/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	230734	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	208509	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	98956	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	48318	26.699	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	53.400%#		
7) Chloroethane-d5	1.648	69	21803m	16.299	ug/L	-0.02
Spiked Amount 50.000	Range 70 - 130		Recovery =	32.600%#		
11) 1,1-Dichloroethene-d2	2.276	65	25364	31.928	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	63.860%		
21) 2-Butanone-d5	4.489	46	138025	84.255	ug/L	0.04
Spiked Amount 100.000	Range 40 - 130		Recovery =	84.250%		
24) Chloroform-d	5.062	84	123104	38.426	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	76.860%		
26) 1,2-Dichloroethane-d4	5.958	65	89250	43.016	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.040%		
32) Benzene-d6	5.970	84	248179	38.567	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	77.140%		
36) 1,2-Dichloropropane-d6	7.312	67	80653	38.946	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	77.900%		
41) Toluene-d8	8.647	98	215583	36.968	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	73.940%#		
43) trans-1,3-Dichloroprop...	8.951	79	34445	32.685	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	65.360%		
47) 2-Hexanone-d5	9.397	63	97909	79.157	ug/L	0.01
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.160%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	116270	39.577	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	79.160%		
66) 1,2-Dichlorobenzene-d4	12.323	152	79404	42.452	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	84.900%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
Data File : VX039305.D
Acq On : 18 Dec 2023 12:45
Operator : JC/MD
Sample : 05806-03ME
Misc : 6.21g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E0LR6ME

Quant Time: Dec 18 21:45:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 18 21:43:41 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/19/2023
Supervised By : Mahesh Dadoda 12/19/2023

