

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
 Data File : VX039339.D
 Acq On : 19 Dec 2023 02:11
 Operator : JC/MD
 Sample : 05835-21 100X
 Misc : 5.33g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B25

Quant Time: Dec 19 04:23:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 18 22:55:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	213455	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	200414	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	96096	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	52200	31.179	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.360%
7) Chloroethane-d5	1.666	69	43837	35.424	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.840%
11) 1,1-Dichloroethene-d2	2.306	65	28715	39.072	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.140%
21) 2-Butanone-d5	4.452	46	148039	97.683	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.680%
24) Chloroform-d	5.056	84	123131	41.545	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.100%
26) 1,2-Dichloroethane-d4	5.958	65	89830	46.801	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.600%
32) Benzene-d6	5.970	84	254914	41.214	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.420%
36) 1,2-Dichloropropane-d6	7.305	67	84268	42.335	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.660%
41) Toluene-d8	8.646	98	224767	40.100	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	8.951	79	36756	36.286	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.580%
47) 2-Hexanone-d5	9.384	63	103570	87.116	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.120%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	124146	43.965	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.940%
66) 1,2-Dichlorobenzene-d4	12.317	152	80476	44.306	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121823\
Data File : VX039339.D
Acq On : 19 Dec 2023 02:11
Operator : JC/MD
Sample : 05835-21 100X
Misc : 5.33g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B25

Quant Time: Dec 19 04:23:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 18 22:55:37 2023
Response via : Initial Calibration

