

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044380.D
 Acq On : 18 Dec 2024 14:01
 Operator : JC/MD
 Sample : P5261-19DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20241210DL

Quant Time: Dec 19 00:45:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115422	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91204	45.067	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.140%
35) Dibromofluoromethane	5.385	113	69475	46.668	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.340%
50) Toluene-d8	8.647	98	254881	50.748	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.079	95	86672	49.387	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.780%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	1029	0.731	ug/l	94
12) 1,1-Dichloroethene	2.306	96	515	0.398	ug/l #	73
44) Trichloroethene	7.123	130	53476	36.703	ug/l	92
64) Tetrachloroethene	9.281	164	752	0.606	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044380.D
Acq On : 18 Dec 2024 14:01
Operator : JC/MD
Sample : P5261-19DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20241210DL

Quant Time: Dec 19 00:45:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

