

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044382.D
 Acq On : 18 Dec 2024 14:48
 Operator : JC/MD
 Sample : P5281-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20241213

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/20/2024
 Supervised By : Mahesh Dadoda 12/20/2024

Quant Time: Dec 19 00:46:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	217882	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94483	44.781	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.560%		
35) Dibromofluoromethane	5.379	113	71615	47.458	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.920%		
50) Toluene-d8	8.647	98	259965	51.063	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.120%		
62) 4-Bromofluorobenzene	11.079	95	82244	46.232	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.460%		
Target Compounds						
16) Acetone	2.380	43	2274	2.644	ug/l #	89
20) Methylene Chloride	2.788	84	534	0.331	ug/l #	82
25) 2-Butanone	4.599	43	2677m	2.189	ug/l	
44) Trichloroethene	7.129	130	584	0.395	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044382.D
Acq On : 18 Dec 2024 14:48
Operator : JC/MD
Sample : P5281-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20241213

Quant Time: Dec 19 00:46:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/20/2024
Supervised By : Mahesh Dadoda 12/20/2024

