

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044385.D
 Acq On : 18 Dec 2024 15:58
 Operator : JC/MD
 Sample : P5280-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-9-20241212

Quant Time: Dec 19 00:48:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97454	44.807	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.620%		
35) Dibromofluoromethane	5.385	113	74309	48.032	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.060%		
50) Toluene-d8	8.647	98	265300	50.830	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.660%		
62) 4-Bromofluorobenzene	11.079	95	86115	47.218	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	94.440%		
Target Compounds						
16) Acetone	2.392	43	1080	1.218	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044385.D
Acq On : 18 Dec 2024 15:58
Operator : JC/MD
Sample : P5280-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-9-20241212

Quant Time: Dec 19 00:48:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

