

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044389.D
 Acq On : 18 Dec 2024 17:32
 Operator : JC/MD
 Sample : P5281-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20241212

Quant Time: Dec 19 00:50:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93938	45.227	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.460%
35) Dibromofluoromethane	5.391	113	70087	46.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	8.647	98	259057	51.231	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.460%
62) 4-Bromofluorobenzene	11.079	95	83204	47.090	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.180%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	737	0.870	ug/l	95
44) Trichloroethene	7.141	130	2669	1.819	ug/l	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044389.D
Acq On : 18 Dec 2024 17:32
Operator : JC/MD
Sample : P5281-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20241212

Quant Time: Dec 19 00:50:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

