

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044402.D
 Acq On : 19 Dec 2024 00:37
 Operator : JC/MD
 Sample : P5317-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20241216

Quant Time: Dec 19 05:07:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	199832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70176	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88339	46.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.800%
35) Dibromofluoromethane	5.385	113	65937	47.642	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.280%
50) Toluene-d8	8.647	98	239168	51.222	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.440%
62) 4-Bromofluorobenzene	11.079	95	75829	46.476	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.960%

Target Compounds						Qvalue
3) Chloromethane	1.294	50	501	0.312	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	13698	10.336	ug/l	98
12) 1,1-Dichloroethene	2.313	96	1397	1.148	ug/l	82
16) Acetone	2.386	43	1776	2.288	ug/l #	86
20) Methylene Chloride	2.788	84	800	0.549	ug/l #	81
27) cis-1,2-Dichloroethene	4.489	96	2032	1.237	ug/l	74
30) Chloroform	5.099	83	968	0.339	ug/l #	57
44) Trichloroethene	7.123	130	694778	512.923	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	465	0.343	ug/l #	58
64) Tetrachloroethene	9.275	164	7782	6.795	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044402.D
Acq On : 19 Dec 2024 00:37
Operator : JC/MD
Sample : P5317-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20241216

Quant Time: Dec 19 05:07:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

