

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044405.D
 Acq On : 19 Dec 2024 01:47
 Operator : JC/MD
 Sample : P5317-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT163S1-20241216

Quant Time: Dec 19 05:08:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	188252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84491	45.001	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.000%		
35) Dibromofluoromethane	5.379	113	62780	48.151	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.300%		
50) Toluene-d8	8.647	98	232616	52.883	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.760%		
62) 4-Bromofluorobenzene	11.079	95	71922	46.793	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.580%		
Target Compounds						
3) Chloromethane	1.294	50	495	0.313	ug/l #	80
16) Acetone	2.380	43	1086	1.419	ug/l #	68
20) Methylene Chloride	2.788	84	506	0.352	ug/l #	79
44) Trichloroethene	7.135	130	1312	1.028	ug/l	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044405.D
Acq On : 19 Dec 2024 01:47
Operator : JC/MD
Sample : P5317-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT163S1-20241216

Quant Time: Dec 19 05:08:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

