

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044414.D
 Acq On : 19 Dec 2024 05:17
 Operator : JC/MD
 Sample : P5317-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D3-20241217

Quant Time: Dec 19 06:41:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	195942	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171336	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85895	45.427	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.860%		
35) Dibromofluoromethane	5.385	113	64519	47.543	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.080%		
50) Toluene-d8	8.647	98	236421	51.638	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.280%		
62) 4-Bromofluorobenzene	11.079	95	71611	44.762	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	89.520%		
Target Compounds						
16) Acetone	2.380	43	826	1.072	ug/l	99
20) Methylene Chloride	2.782	84	433	0.299	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044414.D
Acq On : 19 Dec 2024 05:17
Operator : JC/MD
Sample : P5317-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D3-20241217

Quant Time: Dec 19 06:41:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

