

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
 Data File : VX044416.D
 Acq On : 19 Dec 2024 06:04
 Operator : JC/MD
 Sample : P5317-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20241217

Quant Time: Dec 19 07:13:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	104497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	189474	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	71901	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83114	45.364	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.720%
35) Dibromofluoromethane	5.385	113	63538	48.418	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.647	98	230583	52.083	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.160%
62) 4-Bromofluorobenzene	11.079	95	74919	48.429	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.860%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1947	1.205	ug/l	97
3) Chloromethane	1.294	50	1079	0.698	ug/l #	86
9) 1,1,2-Trichlorotrifluo...	2.319	101	11944	9.366	ug/l	95
12) 1,1-Dichloroethene	2.306	96	3099	2.646	ug/l #	69
16) Acetone	2.392	43	1052	1.408	ug/l #	43
20) Methylene Chloride	2.794	84	368	0.262	ug/l #	59
27) cis-1,2-Dichloroethene	4.495	96	2460	1.556	ug/l	86
30) Chloroform	5.099	83	3176	1.154	ug/l	96
44) Trichloroethene	7.123	130	196003	152.610	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	580	0.451	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121824\
Data File : VX044416.D
Acq On : 19 Dec 2024 06:04
Operator : JC/MD
Sample : P5317-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20241217

Quant Time: Dec 19 07:13:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

