

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121918\
 Data File : VX006654.D
 Acq On : 19 Dec 2018 22:58
 Operator : JC/MD
 Sample : J6199-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 20 04:06:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	633726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1009417	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	967630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	463410	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	350484	51.32	ug/l	0.00
Spiked Amount			Recovery	=	102.64%	
35) Dibromofluoromethane	5.51	113	299751	46.96	ug/l	0.00
Spiked Amount			Recovery	=	93.92%	
50) Toluene-d8	8.71	98	1228944	51.07	ug/l	0.00
Spiked Amount			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	11.13	95	448471	50.88	ug/l	0.00
Spiked Amount			Recovery	=	101.76%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	1.62	94	248	Below Cal	#	53
8) Diethyl Ether	2.19	74	15355	3.801	ug/l	96
13) Acrolein	2.31	56	655	0.637	ug/l #	41
16) Acetone	2.45	43	19695	6.276	ug/l	93
18) Methyl Acetate	2.78	43	23614	2.559	ug/l	97
20) Methylene Chloride	2.86	84	28078	4.327	ug/l	92
25) 2-Butanone	4.70	43	2717	0.616	ug/l #	78
28) Bromochloromethane	5.02	49	297	Below Cal	#	18
36) 1,1-Dichloropropene	5.66	75	41002	4.904	ug/l #	44
38) Carbon Tetrachloride	5.67	117	60047	7.390	ug/l #	15
43) Isopropyl Acetate	6.46	43	29362	2.218	ug/l	99
48) Methyl methacrylate	7.68	41	15389	2.286	ug/l #	11
49) 1,4-Dioxane	7.73	88	132	Below Cal	#	1
51) 4-Methyl-2-Pentanone	8.62	43	155199	18.423	ug/l #	48
52) Toluene	8.79	92	4115	0.251	ug/l	86
54) cis-1,3-Dichloropropene	8.62	75	79625	8.960	ug/l #	63
57) 1,3-Dichloropropane	9.54	76	20741	1.918	ug/l #	43
59) 2-Hexanone	9.54	43	239698	37.189	ug/l #	51
60) Dibromochloromethane	9.41	129	69	2.334	ug/l	72
68) m/p-Xylenes	10.35	106	5835	0.442	ug/l	98
69) o-Xylene	10.69	106	6094	0.468	ug/l	92
76) 1,2,3-Trichloropropane	11.13	75	209833	24.365	ug/l #	40
80) 1,3,5-Trimethylbenzene	11.51	105	14931	0.572	ug/l	98
81) trans-1,4-Dichloro-2-buten	11.13	75	209833	72.105	ug/l #	10
83) tert-Butylbenzene	11.80	119	5504	0.211	ug/l #	47
84) 1,2,4-Trimethylbenzene	11.80	105	43807	1.595	ug/l	97
90) Hexachloroethane	12.58	117	1032	3.512	ug/l #	6
95) Naphthalene	13.83	128	100331	2.949	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121918\
Data File : VX006654.D
Acq On : 19 Dec 2018 22:58
Operator : JC/MD
Sample : J6199-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 20 04:06:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X


ClientSampled :

Tot Ion:168 Resp: 633726




Ion Ratio Lower Upper

168 100

99 52.4 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): VX006654.D (-690) (-)



#5

Bromomethane

Concen: Below Cal

RT: 1.62 min Scan# 76

Delta R.T. -0.02 min

Lab File: VX006654.D


Acq: 19 Dec 2018 22:58

Tgt Ion: 94 Resp: 248



Ion Ratio Lower Upper


94 100

96 136.6 73.4 110.2#



Abundance Ion 93.90 (93.60 to 94.60): VX006654.D (-31) (-)









#8
Diethyl Ether
Concen: 3.801 ug/l
RT: 2.19 min Scan# 170
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 74 Resp: 15355
Ion Ratio Lower Upper
74 100
45 88.7 46.3 138.9



#13
Acrolein
Concen: 0.637 ug/l
RT: 2.31 min Scan# 189
Delta R.T. 0.01 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion: 56 Resp: 655
Ion Ratio Lower Upper
56 100
55 23.2 58.2 87.4#









#16
Acetone
Concen: 6.276 ug/l
RT: 2.45 min Scan# 212
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 19695
Ion Ratio Lower Upper
43 100
58 35.0 25.0 37.6



#18
Methyl Acetate
Concen: 2.559 ug/l
RT: 2.78 min Scan# 267
Delta R.T. 0.01 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion: 43 Resp: 23614
Ion Ratio Lower Upper
43 100
74 24.7 18.6 28.0









#20
Methvlene Chloride
Concen: 4.327 ug/l
RT: 2.86 min Scan# 280
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion:	84	Resp:	28078
Ion	Ratio	Lower	Upper
84	100		
49	103.7	91.8	137.6
51	42.2	28.5	42.7
86	63.8	50.2	75.2



#25
2-Butanone
Concen: 0.616 ug/l
RT: 4.70 min Scan# 581
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion:	43	Resp:	2717
Ion	Ratio	Lower	Upper
43	100		
72	16.5	22.4	33.6#





#28

Bromochloromethane

Concen: Below Cal

RT: 5.02 min Scan# 634





Delta R.T. -0.00 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 49 Resp: 297
Ion Ratio Lower Upper
49 100
129 10.4 0.0 5.4#
130 14.5 77.7 116.5#



#33

1,2-Dichloroethane-d4

Concen: 51.315 ug/l




RT: 6.07 min Scan# 806


Delta R.T. 0.01 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Tgt Ion: 65 Resp: 350484
Ion Ratio Lower Upper
65 100
67 53.8 0.0 105.0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion:114 Resp: 1009417
 Ion Ratio Lower Upper
 114 100
 63 18.4 0.0 34.8
 88 15.7 0.0 28.8



#35
 Dibromofluoromethane
 Concen: 46.958 ug/l
 RT: 5.51 min Scan# 714
 Delta R.T. 0.01 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58

Tgt Ion:113 Resp: 299751
 Ion Ratio Lower Upper
 113 100
 111 101.8 81.6 122.4
 192 20.8 16.7 25.1





#36

1,1-Dichloropropene

Concen: 4.904 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.14 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X

ClientSampleId :





Tot Ion: 75 Resp: 41002

Ion Ratio Lower Upper

75 100

110 5.1 20.2 60.5#

77 0.3 24.8 37.2#



#38

Carbon Tetrachloride

Concen: 7.390 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.12 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58




Tgt Ion: 117 Resp: 60047


Ion Ratio Lower Upper

117 100

119 5.2 79.4 119.2#

121 0.0 24.0 36.0#






#43




Isopropyl Acetate
 Concen: 2.218 ug/l
 RT: 6.46 min Scan# 871
 Delta R.T. 0.01 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58

Instrument :
 MSVOA_X
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
43	100		
61	22.0	16.8	25.2
87	15.7	12.4	18.6




Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 61.00 (60.70 to 61.70): VX0
 Ion 87.00 (86.70 to 87.70): VX0





#48


Methyl methacrylate
 Concen: 2.286 ug/l
 RT: 7.68 min Scan# 1071
 Delta R.T. -0.09 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58

Tgt Ion	Ratio	Lower	Upper
41	100		
69	0.0	72.4	108.6#
39	0.0	45.8	68.8#



Abundance Ion 41.00 (40.70 to 41.70): VX0
 Ion 69.00 (68.70 to 69.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: Below Cal

RT: 7.73 min Scan# 1079

Delta R.T. -0.02 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion: 88 Resp: 132

Ion Ratio Lower Upper

88 100

43 0.0 22.2 33.4#




58 1179.5 51.0 76.4#



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 51.067 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX006654.D


Acq: 19 Dec 2018 22:58

Tgt Ion: 98 Resp: 1228944

Ion Ratio Lower Upper



98 100


100 64.2 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V





#51

4-Methyl-2-Pentanone

Concen: 18.423 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 43 Resp: 155199

Ion Ratio Lower Upper

43 100

58 7.9 31.8 47.6#



#52

Toluene

Concen: 0.251 ug/l

RT: 8.79 min Scan# 1252

Delta R.T. -0.00 min

Lab File: VX006654.D




Acq: 19 Dec 2018 22:58


Tgt Ion: 92 Resp: 4115

Ion Ratio Lower Upper

92 100

91 152.7 137.6 206.4









#54
 cis-1,3-Dichloropropene
 Concen: 8.960 ug/l
 RT: 8.62 min Scan# 1224
 Delta R.T. 0.18 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
75	100		
77	0.4	26.2	39.2#
39	30.6	36.7	55.1#



#57
 1,3-Dichloropropane
 Concen: 1.918 ug/l
 RT: 9.54 min Scan# 1376
 Delta R.T. 0.17 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58

Tgt Ion	Ratio	Lower	Upper
76	100		
78	0.8	26.0	39.0#









#59
2-Hexanone
Concen: 37.189 ug/l
RT: 9.54 min Scan# 1376
Delta R.T. 0.05 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 239698
Ion Ratio Lower Upper
43 100
58 19.8 27.7 83.1#



#60
Dibromochloromethane
Concen: 2.334 ug/l
RT: 9.41 min Scan# 1355
Delta R.T. -0.16 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion: 129 Resp: 69
Ion Ratio Lower Upper
129 100
127 102.9 39.3 117.8









#62
4-Bromofluorobenzene
Concen: 50.883 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 95 Resp: 448471
Ion Ratio Lower Upper
95 100
174 87.2 0.0 182.2
176 84.7 0.0 178.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.01 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion: 117 Resp: 967630
Ion Ratio Lower Upper
117 100
82 49.4 41.0 61.6
119 31.7 25.4 38.0









#68
m/p-Xylenes
Concen: 0.442 ug/l
RT: 10.35 min Scan# 1509
Delta R.T. -0.01 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58




Instrument :
MSVOA_X
ClientSampled :


Tot Ion:106 Resp: 5835
Ion Ratio Lower Upper
106 100
91 191.5 155.8 233.6



#69
o-Xylene
Concen: 0.468 ug/l
RT: 10.69 min Scan# 1565
Delta R.T. -0.01 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion:106 Resp: 6094
Ion Ratio Lower Upper
106 100
91 189.6 101.3 303.7





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:152 Resp: 463410

Ion Ratio Lower Upper

152 100

115 57.1 39.1 117.2




150 158.1 0.0 344.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 24.365 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.16 min

Lab File: VX006654.D


Acq: 19 Dec 2018 22:58

Tgt Ion: 75 Resp: 209833

Ion Ratio Lower Upper



75 100


77 1.3 18.5 55.5#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0









#80
 1,3,5-Trimethylbenzene
 Concen: 0.572 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58




Instrument :
 MSVOA_X
 ClientSampled :


Tot Ion:105 Resp: 14931
 Ion Ratio Lower Upper
 105 100
 120 51.9 25.1 75.3



#81
 trans-1,4-Dichloro-2-butene
 Concen: 72.105 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.06 min
 Lab File: VX006654.D
 Acq: 19 Dec 2018 22:58

Tgt Ion: 75 Resp: 209833
 Ion Ratio Lower Upper
 75 100
 53 0.1 79.7 119.5#
 89 0.0 38.0 57.0#









#83
tert-Butylbenzene
Concen: 0.211 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. 0.04 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Instrument :
MSVOA_X
ClientSampled :

Tot Ion:119 Resp: 5504
Ion Ratio Lower Upper
119 100
91 95.3 27.7 83.1#
134 0.0 11.9 35.5#




Abundance Ion 119.00 (118.70 to 119.70): V
Ion 91.00 (90.70 to 91.70): VX0
Ion 134.00 (133.70 to 134.70): V






#84
1,2,4-Trimethylbenzene
Concen: 1.595 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. -0.00 min
Lab File: VX006654.D
Acq: 19 Dec 2018 22:58

Tgt Ion:105 Resp: 43807
Ion Ratio Lower Upper
105 100
120 48.5 23.2 69.6



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#90

Hexachloroethane

Concen: 3.512 ug/l

RT: 12.58 min Scan# 1875

Delta R.T. -0.01 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:117 Resp: 1032

Ion Ratio Lower Upper




117 100

201 0.0 42.9 128.7#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#95

Naphthalene

Concen: 2.949 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX006654.D

Acq: 19 Dec 2018 22:58


Tgt Ion:128 Resp: 100331

Ion Ratio Lower Upper

128 100

127 12.6 10.1 15.1

129 12.0 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

