

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121918\
 Data File : VX006644.D
 Acq On : 19 Dec 2018 18:31
 Operator : JC/MD
 Sample : J6479-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Dec 20 04:31:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	629413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1012297	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	930800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	438120	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	355324	52.38	ug/l	0.00
Spiked Amount			Recovery	=	104.76%	
35) Dibromofluoromethane	5.51	113	303403	47.40	ug/l	0.00
Spiked Amount			Recovery	=	94.80%	
50) Toluene-d8	8.71	98	1213861	50.30	ug/l	0.00
Spiked Amount			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	11.13	95	435900	49.32	ug/l	0.00
Spiked Amount			Recovery	=	98.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	116676	18.009	ug/l	99
6) Chloroethane	1.73	64	288141	64.230	ug/l	98
11) Tert butyl alcohol	3.06	59	15541	10.266	ug/l #	94
12) 1,1-Dichloroethene	2.38	96	723762	128.700	ug/l	98
20) Methylene Chloride	2.86	84	10339	1.604	ug/l	88
24) 1,1-Dichloroethane	3.70	63	352631	34.111	ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	120567	17.538	ug/l	91
39) Methylcyclohexane	7.46	83	4264	0.386	ug/l	91
40) Benzene	6.15	78	15723	0.617	ug/l	98
42) 1,2-Dichloroethane	6.20	62	53301	6.138	ug/l	98
44) Trichloroethene	7.21	130	8239	1.068	ug/l	98
49) 1,4-Dioxane	7.74	88	313270	1788.943	ug/l	95
69) o-Xylene	10.70	106	5200	0.415	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121918\
Data File : VX006644.D
Acq On : 19 Dec 2018 18:31
Operator : JC/MD
Sample : J6479-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Dec 20 04:31:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

