

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014130.D
 Acq On : 19 Dec 2019 18:57
 Operator : JC/SP
 Sample : K6347-02 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 20 01:01:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	539503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	836339	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	753850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	336022	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	299137	48.92	ug/l	0.00
Spiked Amount			Recovery	=	97.84%	
35) Dibromofluoromethane	5.49	113	245697	48.33	ug/l	0.00
Spiked Amount			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	998499	50.45	ug/l	0.00
Spiked Amount			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	342388	47.32	ug/l	0.00
Spiked Amount			Recovery	=	94.64%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.30	50	1580	0.244	ug/l #	52
8) Diethyl Ether	2.17	74	854	0.215	ug/l	58
10) Methyl Iodide	2.50	142	992	2.761	ug/l #	55
11) Tert butyl alcohol	3.02	59	53287	40.332	ug/l	95
13) Acrolein	2.26	56	1880	2.486	ug/l #	1
16) Acetone	2.44	43	9647	2.430	ug/l #	86
17) Carbon Disulfide	2.56	76	2971	0.953	ug/l #	81
18) Methyl Acetate	2.84	43	7705	0.980	ug/l #	56
20) Methylene Chloride	2.85	84	6084	0.973	ug/l #	91
25) 2-Butanone	4.66	43	1078	0.220	ug/l	98
28) Bromochloromethane	5.01	49	332	0.537	ug/l #	50
31) Cyclohexane	5.57	56	42531	4.655	ug/l	94
36) 1,1-Dichloropropene	5.65	75	36427	4.747	ug/l #	49
38) Carbon Tetrachloride	5.65	117	44775	6.405	ug/l #	19
39) Methylcyclohexane	7.46	83	14374	1.517	ug/l	90
40) Benzene	6.14	78	1561831	66.147	ug/l	99
41) Methacrylonitrile	5.03	41	1046	0.222	ug/l #	45
42) 1,2-Dichloroethane	6.18	62	31534	3.940	ug/l	91
43) Isopropyl Acetate	6.33	43	112595	8.029	ug/l #	63
49) 1,4-Dioxane	7.72	88	207	1.481	ug/l #	1
51) 4-Methyl-2-Pentanone	8.64	43	1773	0.201	ug/l #	82
59) 2-Hexanone	9.50	43	6374	0.899	ug/l	68
69) o-Xylene	10.70	106	84598	8.138	ug/l	96
70) Styrene	10.70	104	5364	0.305	ug/l #	1
73) Isopropylbenzene	11.01	105	6899	0.292	ug/l	99
76) 1,2,3-Trichloropropane	11.13	75	164708	22.578	ug/l #	39
80) 1,3,5-Trimethylbenzene	11.67	105	22084	1.109	ug/l	73
81) trans-1,4-Dichloro-2-buten	11.13	75	164708	63.728	ug/l #	12
95) Naphthalene	13.83	128	10596	0.483	ug/l #	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014130.D
Acq On : 19 Dec 2019 18:57
Operator : JC/SP
Sample : K6347-02 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 20 01:01:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.01 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X


ClientSampled :

Tot Ion:168 Resp: 539503

Ion Ratio Lower Upper

168 100

99 50.3 40.3 60.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

200000

150000



100000

50000

0

Time-->

5.50 5.60 5.70 5.80



#3

Chloromethane

Concen: 0.244 ug/l

RT: 1.30 min Scan# 67

Delta R.T. -0.02 min

Lab File: VX014130.D


Acq: 19 Dec 2019 18:57

Tgt Ion: 50 Resp: 1580

Ion Ratio Lower Upper

50 100

52 5.5 26.2 39.4#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.30

1500


1000


500

0

Time-->

1.26 1.28 1.30 1.32





#8

Diethyl Ether

Concen: 0.215 ug/l

RT: 2.17 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X





ClientSampled :

Tot Ion: 74 Resp: 854

Ion Ratio Lower Upper

74 100

45 55.0 48.1 144.3



#10

Methyl Iodide

Concen: 2.761 ug/l

RT: 2.50 min Scan# 265

Delta R.T. 0.01 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57




Tgt Ion: 142 Resp: 992


Ion Ratio Lower Upper

142 100

127 74.8 31.6 47.4#

141 18.9 11.6 17.4#









#11

Tert butyl alcohol
 Concen: 40.332 ug/l
 RT: 3.02 min Scan# 350
 Delta R.T. -0.02 min
 Lab File: VX014130.D
 Acq: 19 Dec 2019 18:57

Instrument :
 MSVOA_X
 ClientSampleId :




Tot Ion: 59 Resp: 53287
 Ion Ratio Lower Upper
 59 100
 57 8.7 8.4 12.6




#13

Acrolein
 Concen: 2.486 ug/l
 RT: 2.26 min Scan# 225
 Delta R.T. -0.02 min
 Lab File: VX014130.D
 Acq: 19 Dec 2019 18:57

Tgt Ion: 56 Resp: 1880
 Ion Ratio Lower Upper
 56 100
 55 1324.4 56.9 85.3#









#16
Acetone
Concen: 2.430 ug/l
RT: 2.44 min Scan# 254
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 9647
Ion Ratio Lower Upper
43 100
58 38.6 24.9 37.3#



#17
Carbon Disulfide
Concen: 0.953 ug/l
RT: 2.56 min Scan# 275
Delta R.T. 0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion: 76 Resp: 2971
Ion Ratio Lower Upper
76 100
78 2.2 7.2 10.8#









#18
Methyl Acetate
Concen: 0.980 ug/l
RT: 2.84 min Scan# 321
Delta R.T. 0.08 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 7705
Ion Ratio Lower Upper
43 100
74 2.3 19.5 29.3#



#20
Methylene Chloride
Concen: 0.973 ug/l
RT: 2.85 min Scan# 322
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion: 84 Resp: 6084
Ion Ratio Lower Upper
84 100
49 112.0 95.8 143.6
51 28.2 29.8 44.8#
86 55.3 50.8 76.2





#25

2-Butanone

Concen: 0.220 ug/l

RT: 4.66 min Scan# 619

Delta R.T. 0.00 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X





ClientSampled :

Tot Ion: 43 Resp: 1078

Ion Ratio Lower Upper

43 100

72 27.0 21.0 31.4



#28

Bromochloromethane

Concen: 0.537 ug/l

RT: 5.01 min Scan# 676

Delta R.T. 0.01 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57




Tgt Ion: 49 Resp: 332


Ion Ratio Lower Upper

49 100

129 25.3 0.0 5.0#

130 124.4 64.6 97.0#









#31
Cyclohexane
Concen: 4.655 ug/l
RT: 5.57 min Scan# 768
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 56 Resp: 42531
Ion Ratio Lower Upper
56 100
69 23.3 23.2 34.8
84 82.5 69.2 103.8



#33
1,2-Dichloroethane-d4
Concen: 48.922 ug/l
RT: 6.05 min Scan# 847
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion: 65 Resp: 299137
Ion Ratio Lower Upper
65 100
67 53.1 0.0 106.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:114 Resp: 836339

Ion Ratio Lower Upper

114 100

63 20.0 0.0 40.8




88 15.5 0.0 30.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.327 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57


Tgt Ion:113 Resp: 245697

Ion Ratio Lower Upper

113 100

111 102.8 82.0 123.0



192 23.3 19.3 28.9




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.747 ug/l
 RT: 5.65 min Scan# 781
 Delta R.T. -0.14 min
 Lab File: VX014130.D
 Acq: 19 Dec 2019 18:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
75	100		
110	8.3	18.3	54.9#
77	1.3	24.8	37.2#



#38
 Carbon Tetrachloride
 Concen: 6.405 ug/l
 RT: 5.65 min Scan# 782
 Delta R.T. -0.12 min
 Lab File: VX014130.D
 Acq: 19 Dec 2019 18:57

Tgt Ion	Ratio	Lower	Upper
117	100		
119	7.6	76.2	114.4#
121	1.1	23.6	35.4#









#39
Methvlcvclohexane
Concen: 1.517 ug/l
RT: 7.46 min Scan# 1078
Delta R.T. 0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion	83	Resp	14374
Ion Ratio	Lower	Upper	
83	100		
55	65.3	61.0	91.6
98	44.0	38.6	57.8



#40
Benzene
Concen: 66.147 ug/l
RT: 6.14 min Scan# 861
Delta R.T. 0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion	78	Resp	1561831
Ion Ratio	Lower	Upper	
78	100		
77	24.0	18.8	28.2









#41
Methacrylonitrile
Concen: 0.222 ug/l
RT: 5.03 min Scan# 679
Delta R.T. -0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampled :


Tot Ion:	41	Resp:	1046
Ion	Ratio	Lower	Upper
41	100		
39	0.0	44.5	66.7#
67	24.7	57.4	86.0#
52	32.9	23.0	34.4



#42
1,2-Dichloroethane
Concen: 3.940 ug/l
RT: 6.18 min Scan# 869
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion:	62	Resp:	31534
Ion	Ratio	Lower	Upper
62	100		
98	14.0	0.0	21.0





#43

Isopropyl Acetate

Concen: 8.029 ug/l

RT: 6.33 min Scan# 892

Delta R.T. -0.11 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X

ClientSampled :





Tot Ion: 43 Resp: 112595

Ion Ratio Lower Upper

43 100

61 2.4 16.4 24.6#

87 0.3 10.7 16.1#



#49

1,4-Dioxane

Concen: 1.481 ug/l

RT: 7.72 min Scan# 1121

Delta R.T. -0.01 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57




Tgt Ion: 88 Resp: 207


Ion Ratio Lower Upper

88 100

43 174.9 26.5 39.7#

58 0.0 56.8 85.2#









#50
Toluene-d8
Concen: 50.446 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 98 Resp: 998499
Ion Ratio Lower Upper
98 100
100 65.6 52.9 79.3



#51
4-Methyl-2-Pentanone
Concen: 0.201 ug/l
RT: 8.64 min Scan# 1272
Delta R.T. 0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion: 43 Resp: 1773
Ion Ratio Lower Upper
43 100
58 28.9 32.2 48.2#









#59
2-Hexanone
Concen: 0.899 ug/l
RT: 9.50 min Scan# 1413
Delta R.T. 0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 6374
Ion Ratio Lower Upper
43 100
58 32.9 28.0 84.0



#62
4-Bromofluorobenzene
Concen: 47.322 ug/l
RT: 11.13 min Scan# 1680
Delta R.T. -0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion: 95 Resp: 342388
Ion Ratio Lower Upper
95 100
174 86.4 0.0 175.8
176 84.1 0.0 173.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1513

Delta R.T. 0.00 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:117 Resp: 753850

Ion Ratio Lower Upper

117 100

82 53.1 42.2 63.4




119 31.8 25.1 37.7



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#69

o-Xylene

Concen: 8.138 ug/l

RT: 10.70 min Scan# 1609

Delta R.T. 0.00 min

Lab File: VX014130.D


Acq: 19 Dec 2019 18:57

Tgt Ion:106 Resp: 84598

Ion Ratio Lower Upper



106 100


91 215.2 104.2 312.6



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0









#70
Stvrene
Concen: 0.305 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. -0.01 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
104	100		
78	238.0	38.5	57.7#
103	221.0	42.9	64.3#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Tgt Ion	Ratio	Lower	Upper
152	100		
115	55.6	38.3	114.9
150	156.3	0.0	345.4





#73

Isopropylbenzene

Concen: 0.292 ug/l

RT: 11.01 min Scan# 1661

Delta R.T. 0.00 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 6899

Ion Ratio Lower Upper




105 100

120 26.6 13.5 40.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#76

1,2,3-Trichloropropane

Concen: 22.578 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. -0.16 min

Lab File: VX014130.D


Acq: 19 Dec 2019 18:57

Tgt Ion: 75 Resp: 164708

Ion Ratio Lower Upper



75 100


77 1.5 19.3 57.8#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#80

1,3,5-Trimethylbenzene

Concen: 1.109 ug/l

RT: 11.67 min Scan# 1768

Delta R.T. 0.16 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 22084

Ion Ratio Lower Upper




105 100

120 31.7 25.3 75.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 63.728 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. 0.06 min

Lab File: VX014130.D

Acq: 19 Dec 2019 18:57


Tgt Ion: 75 Resp: 164708

Ion Ratio Lower Upper

75 100

53 0.4 76.7 115.1#



89 0.1 34.6 51.8#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0






#95
Naphthalene
Concen: 0.483 ug/l
RT: 13.83 min Scan# 2123
Delta R.T. 0.00 min
Lab File: VX014130.D
Acq: 19 Dec 2019 18:57

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.2	15.4
129	13.3	8.7	13.1



Abundance

Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

