

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014113.D
 Acq On : 19 Dec 2019 11:45
 Operator : JC/SP
 Sample : VX1219WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1219WBL01

Quant Time: Dec 20 01:21:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 13 01:35:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	83068	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.84	114	467675	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	406478	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.05	65	184866	31.07	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.57%
60) 4-Bromofluorobenzene	11.13	95	172411	26.27	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	87.57%
63) Toluene-d8	8.71	98	551133	30.33	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	101.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014113.D
Acq On : 19 Dec 2019 11:45
Operator : JC/SP
Sample : VX1219WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1219WBL01

Quant Time: Dec 20 01:21:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 13 01:35:34 2019
Response via : Initial Calibration

