

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014123.D
 Acq On : 19 Dec 2019 16:14
 Operator : JC/SP
 Sample : K6372-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 20 03:03:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	541647	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	826039	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	713190	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	292149	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	298333	48.60	ug/l	0.00
Spiked Amount			Recovery	=	97.20%	
35) Dibromofluoromethane	5.49	113	245312	48.85	ug/l	0.00
Spiked Amount			Recovery	=	97.70%	
50) Toluene-d8	8.71	98	971006	49.67	ug/l	0.00
Spiked Amount			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	11.13	95	304673	42.63	ug/l	0.00
Spiked Amount			Recovery	=	85.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014123.D
Acq On : 19 Dec 2019 16:14
Operator : JC/SP
Sample : K6372-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 20 03:03:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

