

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014126.D
 Acq On : 19 Dec 2019 17:24
 Operator : JC/SP
 Sample : K6373-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 391218758

Quant Time: Dec 20 03:12:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	552189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	857616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	768494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	339983	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	311242	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
35) Dibromofluoromethane	5.49	113	256036	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	8.71	98	1016051	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	11.14	95	340157	45.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.70%	
Target Compounds						
16) Acetone	2.43	43	69250	17.044	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014126.D
Acq On : 19 Dec 2019 17:24
Operator : JC/SP
Sample : K6373-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
391218758

Quant Time: Dec 20 03:12:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

