

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014149.D
 Acq On : 20 Dec 2019 03:42
 Operator : JC/SP
 Sample : K6372-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Dec 20 05:36:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	560393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	853259	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	762465	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	343627	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	304129	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	
35) Dibromofluoromethane	5.49	113	246552	47.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.06%	
50) Toluene-d8	8.71	98	1011147	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	
62) 4-Bromofluorobenzene	11.14	95	340836	46.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014149.D
Acq On : 20 Dec 2019 03:42
Operator : JC/SP
Sample : K6372-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Dec 20 05:36:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

