

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014150.D
 Acq On : 20 Dec 2019 04:05
 Operator : JC/SP
 Sample : K6372-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Dec 20 05:38:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	552050	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	845401	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	749548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	324492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	302600	48.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.72%	
35) Dibromofluoromethane	5.49	113	245812	47.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.66%	
50) Toluene-d8	8.71	98	996110	49.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.58%	
62) 4-Bromofluorobenzene	11.13	95	327001	44.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.42%	
Target Compounds						
30) Chloroform	5.21	83	10939	1.098	ug/l	89
64) Tetrachloroethene	9.33	164	10983	1.656	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014150.D
Acq On : 20 Dec 2019 04:05
Operator : JC/SP
Sample : K6372-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Dec 20 05:38:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

