

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014151.D
 Acq On : 20 Dec 2019 04:29
 Operator : JC/SP
 Sample : K6372-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Dec 20 05:43:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	556626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	852163	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	764188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	337719	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	308002	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
35) Dibromofluoromethane	5.49	113	253562	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
50) Toluene-d8	8.71	98	1011487	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	11.13	95	340122	46.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.28%	
Target Compounds						
64) Tetrachloroethene	9.34	164	4010	0.593	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014151.D
Acq On : 20 Dec 2019 04:29
Operator : JC/SP
Sample : K6372-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Dec 20 05:43:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

