

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121919\
 Data File : VX014158.D
 Acq On : 20 Dec 2019 07:12
 Operator : JC/SP
 Sample : K6372-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Dec 20 07:55:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	572781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	879746	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	803288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	379194	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	317355	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
35) Dibromofluoromethane	5.49	113	260847	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	8.71	98	1042115	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.13	95	365747	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
Target Compounds						
16) Acetone	2.44	43	6844	1.624	ug/l	95
64) Tetrachloroethene	9.33	164	4468	0.629	ug/l	95
84) 1,2,4-Trimethylbenzene	11.81	105	22661	1.007	ug/l	99
95) Naphthalene	13.83	128	294874	11.915	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121919\
Data File : VX014158.D
Acq On : 20 Dec 2019 07:12
Operator : JC/SP
Sample : K6372-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Dec 20 07:55:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

