

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020181.D
 Acq On : 19 Dec 2020 19:07
 Operator : JC/MD
 Sample : L5096-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 49-IW5678-121220

Manual Integrations
 APPROVED

MMDadoda
 12/21/2020 3:56:40 PM

Quant Time: Dec 21 02:35:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	290305	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.830	114	482458	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	447863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	191080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	191010	49.09	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.18%		
35) Dibromofluoromethane	5.464	113	152950	49.56	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.12%		
50) Toluene-d8	8.696	98	589687	49.64	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.28%		
62) 4-Bromofluorobenzene	11.116	95	204197	44.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.82%		
Target Compounds					Qvalue	
16) Acetone	2.416	43	64868	36.56	ug/l	99
52) Toluene	8.763	92	46388	5.12	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020181.D
Acq On : 19 Dec 2020 19:07
Operator : JC/MD
Sample : L5096-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
49-IW5678-121220

Manual Integrations
APPROVED

MMDadoda
12/21/2020 3:56:40 PM

Quant Time: Dec 21 02:35:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

