

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020182.D
 Acq On : 19 Dec 2020 19:30
 Operator : JC/MD
 Sample : L5096-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 49-FD02-121220

Quant Time: Dec 21 02:36:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	279629	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.830	114	475542	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	442586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	193051	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	187517	50.03	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.06%
35) Dibromofluoromethane	5.465	113	149223	49.05	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.10%
50) Toluene-d8	8.696	98	569066	48.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.20%
62) 4-Bromofluorobenzene	11.122	95	203746	45.46	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.92%
Target Compounds						
16) Acetone	2.416	43	60182	35.21	ug/l	99
52) Toluene	8.763	92	46337	5.19	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020182.D
Acq On : 19 Dec 2020 19:30
Operator : JC/MD
Sample : L5096-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
49-FD02-121220

Quant Time: Dec 21 02:36:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

