

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121920\
 Data File : VX020166.D
 Acq On : 19 Dec 2020 13:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:30:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	125	0.00
2 T	Dichlorodifluoromethane	0.543	0.561	-3.3	138	0.00
3 P	Chloromethane	0.630	0.586	7.0	118	0.00
4 C	Vinyl Chloride	0.686	0.655	4.5#	128	0.00
5 T	Bromomethane	0.220	0.296	-34.5#	177#	0.00
6 T	Chloroethane	0.427	0.413	3.3	120	0.00
7 T	Trichlorofluoromethane	0.882	0.876	0.7	135	0.00
8 T	Diethyl Ether	0.384	0.353	8.1	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.499	0.494	1.0	135	0.00
10 T	Methyl Iodide	0.622	0.639	-2.7	121	0.00
11 T	Tert butyl alcohol	0.150	0.129	14.0	111	0.00
12 CM	1,1-Dichloroethene	0.511	0.492	3.7#	129	0.00
13 T	Acrolein	0.094	0.089	5.3	117	0.00
14 T	Allyl chloride	0.906	0.846	6.6	115	0.00
15 T	Acrylonitrile	0.345	0.320	7.2	109	0.00
16 T	Acetone	0.306	0.306	0.0	124	0.00
17 T	Carbon Disulfide	1.509	1.415	6.2	126	0.00
18 T	Methyl Acetate	0.682	0.626	8.2	109	0.00
19 T	Methyl tert-butyl Ether	1.912	1.802	5.8	108	0.00
20 T	Methylene Chloride	0.689	0.612	11.2	109	0.00
21 T	trans-1,2-Dichloroethene	0.620	0.578	6.8	118	0.00
22 T	Diisopropyl ether	1.850	1.749	5.5	109	0.00
23 T	Vinyl Acetate	1.577	1.518	3.7	107	0.00
24 P	1,1-Dichloroethane	1.090	1.030	5.5	115	0.00
25 T	2-Butanone	0.445	0.429	3.6	113	0.00
26 T	2,2-Dichloropropane	0.899	0.870	3.2	123	0.00
27 T	cis-1,2-Dichloroethene	0.699	0.657	6.0	113	0.00
28 T	Bromochloromethane	0.485	0.445	8.2	111	0.00
29 T	Tetrahydrofuran	0.280	0.264	5.7	109	0.00
30 C	Chloroform	1.109	1.044	5.9#	112	0.00
31 T	Cyclohexane	0.924	0.933	-1.0	136	0.00
32 T	1,1,1-Trichloroethane	0.925	0.898	2.9	124	0.00
33 S	1,2-Dichloroethane-d4	0.670	0.622	7.2	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	114	0.00
35 S	Dibromofluoromethane	0.320	0.319	0.3	113	0.00
36 T	1,1-Dichloropropene	0.494	0.499	-1.0	127	0.00
37 T	Ethyl Acetate	0.514	0.502	2.3	109	0.00
38 T	Carbon Tetrachloride	0.470	0.490	-4.3	128	0.00
39 T	Methylcyclohexane	0.570	0.626	-9.8	143	0.00
40 TM	Benzene	1.538	1.527	0.7	114	0.00
41 T	Methacrylonitrile	0.274	0.277	-1.1	108	0.00
42 TM	1,2-Dichloroethane	0.528	0.518	1.9	109	0.00
43 T	Isopropyl Acetate	0.811	0.802	1.1	108	0.00
44 TM	Trichloroethene	0.398	0.390	2.0	118	0.00
45 C	1,2-Dichloropropane	0.389	0.384	1.3#	111	0.00
46 T	Dibromomethane	0.266	0.257	3.4	108	0.00
47 T	Bromodichloromethane	0.507	0.511	-0.8	110	0.00
48 T	Methyl methacrylate	0.397	0.401	-1.0	108	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020166.D
 Acq On : 19 Dec 2020 13:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:30:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 S	Toluene-d8	1.231	1.244	-1.1	117	0.00
51 T	4-Methyl-2-Pentanone	0.502	0.504	-0.4	108	0.00
52 CM	Toluene	0.939	0.962	-2.4#	117	0.00
53 T	t-1,3-Dichloropropene	0.558	0.573	-2.7	110	0.00
54 T	cis-1,3-Dichloropropene	0.612	0.629	-2.8	110	0.00
55 T	1,1,2-Trichloroethane	0.387	0.388	-0.3	111	0.00
56 T	Ethyl methacrylate	0.567	0.587	-3.5	109	0.00
57 T	1,3-Dichloropropane	0.659	0.658	0.2	110	0.00
58 T	2-Chloroethyl vinyl ether	0.300	0.297	1.0	108	0.00
59 T	2-Hexanone	0.382	0.392	-2.6	111	0.00
60 T	Dibromochloromethane	0.384	0.394	-2.6	110	0.00
61 T	1,2-Dibromoethane	0.404	0.406	-0.5	110	0.00
62 S	4-Bromofluorobenzene	0.471	0.467	0.8	115	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
64 T	Tetrachloroethene	0.377	0.385	-2.1	127	0.00
65 PM	Chlorobenzene	1.073	1.058	1.4	112	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.389	-1.0	112	0.00
67 C	Ethyl Benzene	1.875	1.933	-3.1#	118	0.00
68 T	m/p-Xylenes	0.700	0.735	-5.0	118	0.00
69 T	o-Xylene	0.674	0.692	-2.7	113	0.00
70 T	Styrene	1.139	1.188	-4.3	112	0.00
71 P	Bromoform	0.298	0.303	-1.7	112	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.609	3.826	-6.0	123	0.00
74 T	N-aryl acetate	1.455	1.472	-1.2	107	0.00
75 P	1,1,2,2-Tetrachloroethane	1.310	1.261	3.7	108	0.00
76 T	1,2,3-Trichloropropane	1.232	1.216	1.3	112	0.00
77 T	Bromobenzene	0.954	0.919	3.7	112	0.00
78 T	n-propylbenzene	4.159	4.426	-6.4	123	0.00
79 T	2-Chlorotoluene	2.593	2.628	-1.3	115	0.00
80 T	1,3,5-Trimethylbenzene	3.059	3.223	-5.4	119	0.00
81 T	trans-1,4-Dichloro-2-butene	0.415	0.418	-0.7	108	0.00
82 T	4-Chlorotoluene	3.029	3.097	-2.2	115	0.00
83 T	tert-Butylbenzene	2.935	3.121	-6.3	123	0.00
84 T	1,2,4-Trimethylbenzene	3.105	3.255	-4.8	115	0.00
85 T	sec-Butylbenzene	3.387	3.689	-8.9	130	0.00
86 T	p-Isopropyltoluene	3.124	3.413	-9.3	126	0.00
87 T	1,3-Dichlorobenzene	1.705	1.653	3.0	110	0.00
88 T	1,4-Dichlorobenzene	1.779	1.692	4.9	111	0.00
89 T	n-Butylbenzene	2.737	3.053	-11.5	132	0.00
90 T	Hexachloroethane	0.539	0.555	-3.0	123	0.00
91 T	1,2-Dichlorobenzene	1.697	1.681	0.9	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.275	0.279	-1.5	109	0.00
93 T	1,2,4-Trichlorobenzene	1.091	1.156	-6.0	122	0.00
94 T	Hexachlorobutadiene	0.484	0.509	-5.2	149	0.00
95 T	Naphthalene	3.456	3.696	-6.9	114	0.00
96 T	1,2,3-Trichlorobenzene	1.080	1.107	-2.5	118	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020166.D
Acq On : 19 Dec 2020 13:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 02:30:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6