

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121920\
 Data File : VX020192.D
 Acq On : 19 Dec 2020 23:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:39:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	50.000	50.624	-1.2	124	0.00
3 P	Chloromethane	50.000	46.940	6.1	109	0.00
4 C	Vinyl Chloride	50.000	49.551	0.9#	121	0.00
5 T	Bromomethane	50.000	50.408	-0.8	122	0.00
6 T	Chloroethane	50.000	49.588	0.8	113	0.00
7 T	Trichlorofluoromethane	50.000	50.643	-1.3	126	0.00
8 T	Diethyl Ether	50.000	48.650	2.7	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.997	2.0	123	0.00
10 T	Methyl Iodide	50.000	49.972	0.1	117	0.00
11 T	Tert butyl alcohol	250.000	237.676	4.9	113	0.00
12 CM	1,1-Dichloroethene	50.000	49.425	1.2#	121	0.00
13 T	Acrolein	250.000	245.408	1.8	111	0.00
14 T	Allyl chloride	50.000	48.367	3.3	109	0.00
15 T	Acrylonitrile	250.000	249.384	0.2	107	0.00
16 T	Acetone	250.000	221.371	11.5	101	0.00
17 T	Carbon Disulfide	50.000	48.783	2.4	120	0.00
18 T	Methyl Acetate	50.000	50.415	-0.8	109	0.00
19 T	Methyl tert-butyl Ether	50.000	49.743	0.5	104	0.00
20 T	Methylene Chloride	50.000	46.658	6.7	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.368	3.3	112	0.00
22 T	Diisopropyl ether	50.000	50.322	-0.6	106	0.00
23 T	Vinyl Acetate	250.000	255.996	-2.4	105	0.00
24 P	1,1-Dichloroethane	50.000	49.787	0.4	111	0.00
25 T	2-Butanone	250.000	245.906	1.6	106	0.00
26 T	2,2-Dichloropropane	50.000	43.785	12.4	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.224	1.6	108	0.00
28 T	Bromochloromethane	50.000	48.421	3.2	107	0.00
29 T	Tetrahydrofuran	250.000	255.212	-2.1	108	0.00
30 C	Chloroform	50.000	49.701	0.6#	109	0.00
31 T	Cyclohexane	50.000	51.121	-2.2	126	0.00
32 T	1,1,1-Trichloroethane	50.000	50.932	-1.9	119	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.059	1.9	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	112	0.00
35 S	Dibromofluoromethane	50.000	49.093	1.8	109	0.00
36 T	1,1-Dichloropropene	50.000	48.808	2.4	120	0.00
37 T	Ethyl Acetate	50.000	49.361	1.3	107	0.00
38 T	Carbon Tetrachloride	50.000	50.152	-0.3	121	0.00
39 T	Methylcyclohexane	50.000	49.331	1.3	126	0.00
40 TM	Benzene	50.000	48.833	2.3	110	0.00
41 T	Methacrylonitrile	50.000	49.937	0.1	104	0.00
42 TM	1,2-Dichloroethane	50.000	48.600	2.8	105	0.00
43 T	Isopropyl Acetate	50.000	49.506	1.0	106	0.00
44 TM	Trichloroethene	50.000	47.597	4.8	112	0.00
45 C	1,2-Dichloropropane	50.000	49.034	1.9#	108	0.00
46 T	Dibromomethane	50.000	47.294	5.4	103	0.00
47 T	Bromodichloromethane	50.000	50.010	-0.0	107	0.00
48 T	Methyl methacrylate	50.000	50.880	-1.8	107	0.00
49 T	1,4-Dioxane	1000.000	1016.075	-1.6	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020192.D
 Acq On : 19 Dec 2020 23:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:39:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.114	1.8	111	0.00
51 T	4-Methyl-2-Pentanone	250.000	256.116	-2.4	108	0.00
52 CM	Toluene	50.000	49.587	0.8#	110	0.00
53 T	t-1,3-Dichloropropene	50.000	49.287	1.4	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.243	1.5	104	0.00
55 T	1,1,2-Trichloroethane	50.000	48.670	2.7	105	0.00
56 T	Ethyl methacrylate	50.000	51.869	-3.7	107	0.00
57 T	1,3-Dichloropropane	50.000	48.928	2.1	105	0.00
58 T	2-Chloroethyl vinyl ether	250.000	240.671	3.7	102	0.00
59 T	2-Hexanone	250.000	255.433	-2.2	108	0.00
60 T	Dibromochloromethane	50.000	50.730	-1.5	107	0.00
61 T	1,2-Dibromoethane	50.000	49.612	0.8	106	0.00
62 S	4-Bromofluorobenzene	50.000	48.723	2.6	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	49.617	0.8	119	0.00
65 PM	Chlorobenzene	50.000	49.149	1.7	108	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.235	-0.5	108	0.00
67 C	Ethyl Benzene	50.000	50.867	-1.7#	112	0.00
68 T	m/p-Xylenes	100.000	103.171	-3.2	112	0.00
69 T	o-Xylene	50.000	50.556	-1.1	108	0.00
70 T	Styrene	50.000	51.979	-4.0	108	0.00
71 P	Bromoform	50.000	50.525	-1.0	108	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73 T	Isopropylbenzene	50.000	51.274	-2.5	114	0.00
74 T	N-aryl acetate	50.000	52.268	-4.5	106	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.317	1.4	106	0.00
76 T	1,2,3-Trichloropropane	50.000	49.149	1.7	107	0.00
77 T	Bromobenzene	50.000	48.178	3.6	108	0.00
78 T	n-propylbenzene	50.000	51.477	-3.0	115	0.00
79 T	2-Chlorotoluene	50.000	49.626	0.7	109	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.954	-3.9	113	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.955	4.1	99	0.00
82 T	4-Chlorotoluene	50.000	50.390	-0.8	109	0.00
83 T	tert-Butylbenzene	50.000	52.290	-4.6	117	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.882	-3.8	110	0.00
85 T	sec-Butylbenzene	50.000	53.099	-6.2	122	0.00
86 T	p-Isopropyltoluene	50.000	52.045	-4.1	115	0.00
87 T	1,3-Dichlorobenzene	50.000	47.786	4.4	104	0.00
88 T	1,4-Dichlorobenzene	50.000	47.300	5.4	106	0.00
89 T	n-Butylbenzene	50.000	51.741	-3.5	118	0.00
90 T	Hexachloroethane	50.000	51.383	-2.8	118	0.00
91 T	1,2-Dichlorobenzene	50.000	48.427	3.1	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.138	-2.3	106	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.334	-2.7	114	0.00
94 T	Hexachlorobutadiene	50.000	50.598	-1.2	131	0.00
95 T	Naphthalene	50.000	54.498	-9.0	111	0.00
96 T	1,2,3-Trichlorobenzene	50.000	50.699	-1.4	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020192.D
Acq On : 19 Dec 2020 23:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 02:39:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6