

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121920\
 Data File : VX020194.D
 Acq On : 20 Dec 2020 00:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:40:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	50.000	51.008	-2.0	122	0.00
3 P	Chloromethane	50.000	46.960	6.1	107	0.00
4 C	Vinyl Chloride	50.000	49.497	1.0#	119	0.00
5 T	Bromomethane	50.000	51.994	-4.0	123	0.00
6 T	Chloroethane	50.000	49.682	0.6	111	0.00
7 T	Trichlorofluoromethane	50.000	50.299	-0.6	123	0.00
8 T	Diethyl Ether	50.000	48.515	3.0	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.083	3.8	118	0.00
10 T	Methyl Iodide	50.000	47.868	4.3	109	0.00
11 T	Tert butyl alcohol	250.000	223.360	10.7	104	0.00
12 CM	1,1-Dichloroethene	50.000	48.940	2.1#	118	0.00
13 T	Acrolein	250.000	265.437	-6.2	118	0.00
14 T	Allyl chloride	50.000	47.499	5.0	105	0.00
15 T	Acrylonitrile	250.000	246.428	1.4	104	0.00
16 T	Acetone	250.000	217.674	12.9	97	0.00
17 T	Carbon Disulfide	50.000	47.532	4.9	115	0.00
18 T	Methyl Acetate	50.000	50.010	-0.0	107	0.00
19 T	Methyl tert-butyl Ether	50.000	49.543	0.9	102	0.00
20 T	Methylene Chloride	50.000	46.389	7.2	103	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.913	4.2	109	0.00
22 T	Diisopropyl ether	50.000	49.896	0.2	104	0.00
23 T	Vinyl Acetate	250.000	251.875	-0.8	101	0.00
24 P	1,1-Dichloroethane	50.000	49.364	1.3	108	0.00
25 T	2-Butanone	250.000	240.270	3.9	102	0.00
26 T	2,2-Dichloropropane	50.000	42.208	15.6	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.433	3.1	105	0.00
28 T	Bromochloromethane	50.000	49.831	0.3	108	0.00
29 T	Tetrahydrofuran	250.000	249.990	0.0	104	0.00
30 C	Chloroform	50.000	49.411	1.2#	106	0.00
31 T	Cyclohexane	50.000	50.347	-0.7	122	0.01
32 T	1,1,1-Trichloroethane	50.000	50.195	-0.4	115	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.356	-0.7	108	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	109	0.00
35 S	Dibromofluoromethane	50.000	50.678	-1.4	110	0.00
36 T	1,1-Dichloropropene	50.000	48.649	2.7	117	0.00
37 T	Ethyl Acetate	50.000	49.126	1.7	104	0.00
38 T	Carbon Tetrachloride	50.000	50.258	-0.5	118	0.01
39 T	Methylcyclohexane	50.000	48.587	2.8	121	0.00
40 TM	Benzene	50.000	48.556	2.9	107	0.00
41 T	Methacrylonitrile	50.000	50.162	-0.3	102	0.00
42 TM	1,2-Dichloroethane	50.000	48.831	2.3	103	0.00
43 T	Isopropyl Acetate	50.000	48.748	2.5	102	0.00
44 TM	Trichloroethene	50.000	47.264	5.5	109	0.00
45 C	1,2-Dichloropropane	50.000	49.385	1.2#	106	0.00
46 T	Dibromomethane	50.000	47.878	4.2	102	0.00
47 T	Bromodichloromethane	50.000	49.709	0.6	104	0.00
48 T	Methyl methacrylate	50.000	50.105	-0.2	103	0.00
49 T	1,4-Dioxane	1000.000	906.044	9.4	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020194.D
 Acq On : 20 Dec 2020 00:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:40:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	50.880	-1.8	113	0.00
51 T	4-Methyl-2-Pentanone	250.000	250.219	-0.1	103	0.00
52 CM	Toluene	50.000	49.345	1.3#	107	0.00
53 T	t-1,3-Dichloropropene	50.000	47.651	4.7	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.490	3.0	99	0.00
55 T	1,1,2-Trichloroethane	50.000	49.036	1.9	103	0.00
56 T	Ethyl methacrylate	50.000	50.624	-1.2	102	0.00
57 T	1,3-Dichloropropane	50.000	49.100	1.8	103	0.00
58 T	2-Chloroethyl vinyl ether	250.000	242.021	3.2	100	0.00
59 T	2-Hexanone	250.000	246.909	1.2	102	0.00
60 T	Dibromochloromethane	50.000	49.377	1.2	101	0.00
61 T	1,2-Dibromoethane	50.000	49.071	1.9	103	0.00
62 S	4-Bromofluorobenzene	50.000	50.005	-0.0	111	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	48.989	2.0	115	0.00
65 PM	Chlorobenzene	50.000	47.984	4.0	104	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.381	1.2	104	0.00
67 C	Ethyl Benzene	50.000	49.670	0.7#	108	0.00
68 T	m/p-Xylenes	100.000	99.430	0.6	106	0.00
69 T	o-Xylene	50.000	49.536	0.9	104	0.00
70 T	Styrene	50.000	50.790	-1.6	103	0.00
71 P	Bromoform	50.000	48.143	3.7	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	51.040	-2.1	111	0.00
74 T	N-aryl acetate	50.000	50.142	-0.3	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.278	1.4	104	0.00
76 T	1,2,3-Trichloropropane	50.000	48.395	3.2	103	0.00
77 T	Bromobenzene	50.000	47.013	6.0	103	0.00
78 T	n-propylbenzene	50.000	50.839	-1.7	111	0.00
79 T	2-Chlorotoluene	50.000	49.490	1.0	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.614	-3.2	109	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.536	6.9	94	0.00
82 T	4-Chlorotoluene	50.000	49.126	1.7	104	0.00
83 T	tert-Butylbenzene	50.000	50.199	-0.4	109	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.270	-2.5	106	0.00
85 T	sec-Butylbenzene	50.000	51.970	-3.9	117	0.00
86 T	p-Isopropyltoluene	50.000	51.184	-2.4	111	0.00
87 T	1,3-Dichlorobenzene	50.000	47.846	4.3	102	0.00
88 T	1,4-Dichlorobenzene	50.000	45.498	9.0	100	0.00
89 T	n-Butylbenzene	50.000	51.013	-2.0	113	0.00
90 T	Hexachloroethane	50.000	50.421	-0.8	113	0.00
91 T	1,2-Dichlorobenzene	50.000	47.607	4.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.262	1.5	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.504	3.0	105	0.00
94 T	Hexachlorobutadiene	50.000	49.906	0.2	126	0.00
95 T	Naphthalene	50.000	53.086	-6.2	106	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.447	3.1	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020194.D
Acq On : 20 Dec 2020 00:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 02:40:25 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6