

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020216.D
 Acq On : 20 Dec 2020 09:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:48:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	111	0.00
2 T	Dichlorodifluoromethane	50.000	52.615	-5.2	124	0.00
3 P	Chloromethane	50.000	48.701	2.6	109	0.00
4 C	Vinyl Chloride	50.000	50.433	-0.9#	119	0.00
5 T	Bromomethane	50.000	50.577	-1.2	118	0.00
6 T	Chloroethane	50.000	51.490	-3.0	113	0.00
7 T	Trichlorofluoromethane	50.000	51.968	-3.9	125	0.00
8 T	Diethyl Ether	50.000	49.420	1.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.870	0.3	121	0.00
10 T	Methyl Iodide	50.000	48.724	2.6	110	0.00
11 T	Tert butyl alcohol	250.000	244.620	2.2	112	0.00
12 CM	1,1-Dichloroethene	50.000	50.512	-1.0#	120	0.00
13 T	Acrolein	250.000	233.810	6.5	102	0.00
14 T	Allyl chloride	50.000	47.850	4.3	104	0.00
15 T	Acrylonitrile	250.000	256.108	-2.4	106	0.00
16 T	Acetone	250.000	226.351	9.5	100	0.00
17 T	Carbon Disulfide	50.000	49.344	1.3	117	0.00
18 T	Methyl Acetate	50.000	52.668	-5.3	111	0.00
19 T	Methyl tert-butyl Ether	50.000	50.455	-0.9	102	0.00
20 T	Methylene Chloride	50.000	47.899	4.2	104	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.875	0.3	112	0.00
22 T	Diisopropyl ether	50.000	51.468	-2.9	105	0.00
23 T	Vinyl Acetate	250.000	256.697	-2.7	101	0.00
24 P	1,1-Dichloroethane	50.000	50.879	-1.8	110	0.00
25 T	2-Butanone	250.000	249.829	0.1	104	0.00
26 T	2,2-Dichloropropane	50.000	36.814	26.4#	83	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.998	0.0	106	0.00
28 T	Bromochloromethane	50.000	49.669	0.7	106	0.00
29 T	Tetrahydrofuran	250.000	260.783	-4.3	107	0.00
30 C	Chloroform	50.000	51.151	-2.3#	108	0.00
31 T	Cyclohexane	50.000	52.626	-5.3	125	0.00
32 T	1,1,1-Trichloroethane	50.000	52.452	-4.9	119	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.895	0.2	106	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	49.245	1.5	106	0.00
36 T	1,1-Dichloropropene	50.000	49.647	0.7	118	0.00
37 T	Ethyl Acetate	50.000	49.636	0.7	105	0.00
38 T	Carbon Tetrachloride	50.000	51.207	-2.4	120	0.00
39 T	Methylcyclohexane	50.000	50.003	-0.0	123	0.00
40 TM	Benzene	50.000	49.728	0.5	109	0.00
41 T	Methacrylonitrile	50.000	50.610	-1.2	102	0.00
42 TM	1,2-Dichloroethane	50.000	49.582	0.8	104	0.00
43 T	Isopropyl Acetate	50.000	50.180	-0.4	104	0.00
44 TM	Trichloroethene	50.000	48.161	3.7	110	0.00
45 C	1,2-Dichloropropane	50.000	50.079	-0.2#	107	0.00
46 T	Dibromomethane	50.000	48.566	2.9	103	0.00
47 T	Bromodichloromethane	50.000	51.064	-2.1	106	0.00
48 T	Methyl methacrylate	50.000	51.486	-3.0	105	0.00
49 T	1,4-Dioxane	1000.000	966.757	3.3	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121920\
 Data File : VX020216.D
 Acq On : 20 Dec 2020 09:25
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:48:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.571	0.9	109	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.072	-3.6	106	0.00
52 CM	Toluene	50.000	50.503	-1.0#	109	0.00
53 T	t-1,3-Dichloropropene	50.000	48.338	3.3	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.609	2.8	99	0.00
55 T	1,1,2-Trichloroethane	50.000	50.057	-0.1	105	0.00
56 T	Ethyl methacrylate	50.000	52.730	-5.5	105	0.00
57 T	1,3-Dichloropropane	50.000	49.950	0.1	104	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.728	2.1	101	0.00
59 T	2-Hexanone	250.000	257.948	-3.2	106	0.00
60 T	Dibromochloromethane	50.000	51.205	-2.4	105	0.00
61 T	1,2-Dibromoethane	50.000	50.302	-0.6	104	0.00
62 S	4-Bromofluorobenzene	50.000	49.005	2.0	108	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	47.817	4.4	113	0.00
65 PM	Chlorobenzene	50.000	48.771	2.5	105	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.749	-1.5	107	0.00
67 C	Ethyl Benzene	50.000	50.995	-2.0#	110	0.00
68 T	m/p-Xylenes	100.000	102.190	-2.2	109	0.00
69 T	o-Xylene	50.000	51.015	-2.0	107	0.00
70 T	Styrene	50.000	52.192	-4.4	106	0.00
71 P	Bromoform	50.000	50.043	-0.1	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	51.268	-2.5	113	0.00
74 T	N-aryl acetate	50.000	51.274	-2.5	103	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.222	-0.4	107	0.00
76 T	1,2,3-Trichloropropane	50.000	48.761	2.5	105	0.00
77 T	Bromobenzene	50.000	48.513	3.0	108	0.00
78 T	n-propylbenzene	50.000	50.943	-1.9	112	0.00
79 T	2-Chlorotoluene	50.000	50.316	-0.6	109	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.593	-3.2	111	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	44.872	10.3	92	0.00
82 T	4-Chlorotoluene	50.000	50.016	-0.0	107	0.00
83 T	tert-Butylbenzene	50.000	51.398	-2.8	114	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.767	-3.5	108	0.00
85 T	sec-Butylbenzene	50.000	52.400	-4.8	119	0.00
86 T	p-Isopropyltoluene	50.000	52.045	-4.1	114	0.00
87 T	1,3-Dichlorobenzene	50.000	49.032	1.9	106	0.00
88 T	1,4-Dichlorobenzene	50.000	47.018	6.0	105	0.00
89 T	n-Butylbenzene	50.000	51.374	-2.7	115	0.00
90 T	Hexachloroethane	50.000	51.067	-2.1	116	0.00
91 T	1,2-Dichlorobenzene	50.000	47.189	5.6	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.458	-0.9	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.038	1.9	107	0.00
94 T	Hexachlorobutadiene	50.000	51.030	-2.1	131	0.00
95 T	Naphthalene	50.000	53.874	-7.7	109	0.00
96 T	1,2,3-Trichlorobenzene	50.000	50.743	-1.5	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020216.D
Acq On : 20 Dec 2020 09:25
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 02:48:08 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820w.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6