

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX121920\
 Data File : VX020218.D
 Acq On : 20 Dec 2020 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:48:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	112	0.01
2 T	Dichlorodifluoromethane	50.000	53.070	-6.1	127	0.00
3 P	Chloromethane	50.000	48.482	3.0	110	0.00
4 C	Vinyl Chloride	50.000	50.258	-0.5#	120	0.00
5 T	Bromomethane	50.000	52.771	-5.5	124	0.00
6 T	Chloroethane	50.000	50.608	-1.2	113	0.00
7 T	Trichlorofluoromethane	50.000	51.616	-3.2	126	0.00
8 T	Diethyl Ether	50.000	48.472	3.1	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.883	2.2	120	0.00
10 T	Methyl Iodide	50.000	46.277	7.4	105	0.00
11 T	Tert butyl alcohol	250.000	233.216	6.7	108	0.00
12 CM	1,1-Dichloroethene	50.000	49.824	0.4#	119	0.00
13 T	Acrolein	250.000	252.415	-1.0	112	0.00
14 T	Allyl chloride	50.000	47.382	5.2	104	0.00
15 T	Acrylonitrile	250.000	253.534	-1.4	106	0.00
16 T	Acetone	250.000	228.841	8.5	102	0.00
17 T	Carbon Disulfide	50.000	48.574	2.9	117	0.00
18 T	Methyl Acetate	50.000	53.275	-6.5	113	0.00
19 T	Methyl tert-butyl Ether	50.000	50.137	-0.3	103	0.00
20 T	Methylene Chloride	50.000	47.483	5.0	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.392	1.2	112	0.00
22 T	Diisopropyl ether	50.000	51.056	-2.1	105	0.00
23 T	Vinyl Acetate	250.000	253.200	-1.3	101	0.00
24 P	1,1-Dichloroethane	50.000	50.024	-0.0	109	0.00
25 T	2-Butanone	250.000	247.332	1.1	104	0.00
26 T	2,2-Dichloropropane	50.000	35.743	28.5#	81	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.323	1.4	106	0.00
28 T	Bromochloromethane	50.000	49.956	0.1	108	0.00
29 T	Tetrahydrofuran	250.000	257.277	-2.9	107	0.00
30 C	Chloroform	50.000	50.626	-1.3#	108	0.00
31 T	Cyclohexane	50.000	51.085	-2.2	123	0.00
32 T	1,1,1-Trichloroethane	50.000	51.159	-2.3	117	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.875	0.3	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	50.323	-0.6	108	0.00
36 T	1,1-Dichloropropene	50.000	49.570	0.9	118	0.00
37 T	Ethyl Acetate	50.000	49.700	0.6	104	0.00
38 T	Carbon Tetrachloride	50.000	50.931	-1.9	119	0.00
39 T	Methylcyclohexane	50.000	48.907	2.2	120	0.00
40 TM	Benzene	50.000	49.731	0.5	108	0.00
41 T	Methacrylonitrile	50.000	50.954	-1.9	103	0.00
42 TM	1,2-Dichloroethane	50.000	49.645	0.7	104	0.00
43 T	Isopropyl Acetate	50.000	49.377	1.2	102	0.00
44 TM	Trichloroethene	50.000	47.823	4.4	109	0.00
45 C	1,2-Dichloropropane	50.000	50.468	-0.9#	107	0.00
46 T	Dibromomethane	50.000	48.418	3.2	103	0.00
47 T	Bromodichloromethane	50.000	50.868	-1.7	105	0.00
48 T	Methyl methacrylate	50.000	51.495	-3.0	105	0.00
49 T	1,4-Dioxane	1000.000	945.214	5.5	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
 Data File : VX020218.D
 Acq On : 20 Dec 2020 10:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 02:48:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.861	0.3	109	0.00
51 T	4-Methyl-2-Pentanone	250.000	258.132	-3.3	106	0.00
52 CM	Toluene	50.000	50.270	-0.5#	108	0.00
53 T	t-1,3-Dichloropropene	50.000	47.880	4.2	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.342	3.3	98	0.00
55 T	1,1,2-Trichloroethane	50.000	49.585	0.8	104	0.00
56 T	Ethyl methacrylate	50.000	52.054	-4.1	104	0.00
57 T	1,3-Dichloropropane	50.000	50.167	-0.3	104	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	245.045	2.0	101	0.00
59 T	2-Hexanone	250.000	255.556	-2.2	105	0.00
60 T	Dibromochloromethane	50.000	50.153	-0.3	102	0.00
61 T	1,2-Dibromoethane	50.000	50.017	-0.0	104	0.00
62 S	4-Bromofluorobenzene	50.000	50.216	-0.4	110	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	106	0.00
64 T	Tetrachloroethene	50.000	49.235	1.5	116	0.00
65 PM	Chlorobenzene	50.000	48.212	3.6	104	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.520	-1.0	107	0.00
67 C	Ethyl Benzene	50.000	50.199	-0.4#	109	0.00
68 T	m/p-Xylenes	100.000	101.463	-1.5	108	0.00
69 T	o-Xylene	50.000	50.133	-0.3	105	0.00
70 T	Styrene	50.000	51.764	-3.5	106	0.00
71 P	Bromoform	50.000	50.068	-0.1	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	50.952	-1.9	111	0.00
74 T	N-aryl acetate	50.000	50.959	-1.9	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.727	0.5	105	0.00
76 T	1,2,3-Trichloropropane	50.000	49.190	1.6	105	0.00
77 T	Bromobenzene	50.000	47.992	4.0	105	0.00
78 T	n-propylbenzene	50.000	51.382	-2.8	112	0.00
79 T	2-Chlorotoluene	50.000	49.875	0.3	107	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.630	-3.3	110	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.822	12.4	89	0.00
82 T	4-Chlorotoluene	50.000	50.325	-0.7	107	0.00
83 T	tert-Butylbenzene	50.000	51.142	-2.3	112	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.517	-3.0	107	0.00
85 T	sec-Butylbenzene	50.000	52.528	-5.1	118	0.00
86 T	p-Isopropyltoluene	50.000	52.033	-4.1	113	0.00
87 T	1,3-Dichlorobenzene	50.000	47.686	4.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	47.092	5.8	104	0.00
89 T	n-Butylbenzene	50.000	51.118	-2.2	114	0.00
90 T	Hexachloroethane	50.000	50.280	-0.6	113	0.00
91 T	1,2-Dichlorobenzene	50.000	48.593	2.8	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.195	-0.4	102	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.719	0.6	108	0.00
94 T	Hexachlorobutadiene	50.000	51.166	-2.3	130	0.00
95 T	Naphthalene	50.000	53.597	-7.2	107	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.693	0.6	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121920\
Data File : VX020218.D
Acq On : 20 Dec 2020 10:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 02:48:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6