

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121921\
 Data File : VX025906.D
 Acq On : 18 Dec 2021 13:05
 Operator : JC/MD
 Sample : M5079-05
 Misc : 4.21g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 338

Quant Time: Dec 19 05:55:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	252893	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	214245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88129	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97906	45.19	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.38%
35) Dibromofluoromethane	5.397	113	79030	44.81	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.62%
50) Toluene-d8	8.653	98	292693	45.55	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.10%#
62) 4-Bromofluorobenzene	11.085	95	95678	39.20	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.40%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121921\
Data File : VX025906.D
Acq On : 18 Dec 2021 13:05
Operator : JC/MD
Sample : M5079-05
Misc : 4.21g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
338

Quant Time: Dec 19 05:55:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

