



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121921\
 Data File : VX025901.D
 Acq On : 18 Dec 2021 09:52
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Title : SW846 8260
 Last Update : Tue Dec 14 07:02:11 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	6211	PASS
75	95	30	60	52.9	15880	PASS
95	95	100	100	100.0	30027	PASS
96	95	5	9	7.2	2165	PASS
173	174	0.00	2	1.1	224	PASS
174	95	50	100	68.4	20530	PASS
175	174	5	9	7.7	1589	PASS
176	174	95	101	96.7	19847	PASS
177	176	5	9	6.9	1377	PASS