

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033476.D
 Acq On : 19 Dec 2022 12:04
 Operator : JC/MD
 Sample : PB149705TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149705TB

Quant Time: Dec 20 04:59:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184740	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36507	48.567	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.140%
35) Dibromofluoromethane	5.385	113	40840	49.150	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	8.647	98	86521	49.775	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	67719	50.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	9927	17.847	ug/l	99
20) Methylene Chloride	2.794	84	7575	6.106	ug/l	88
25) 2-Butanone	4.568	43	5383	6.742	ug/l	94
37) Ethyl Acetate	4.721	43	15881	10.370	ug/l	99
43) Isopropyl Acetate	6.336	43	68692	27.878	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.251	63	1382	1.671	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033476.D
Acq On : 19 Dec 2022 12:04
Operator : JC/MD
Sample : PB149705TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149705TB

Quant Time: Dec 20 04:59:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

