

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033478.D
 Acq On : 19 Dec 2022 12:50
 Operator : JC/MD
 Sample : PB149705ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149705ZHE#02

Quant Time: Dec 20 05:00:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	24980	35.094	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	70.180%#	
35) Dibromofluoromethane	5.391	113	28493	35.123	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	70.240%#	
50) Toluene-d8	8.653	98	57810	34.065	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	68.140%#	
62) 4-Bromofluorobenzene	11.079	95	45037	34.708	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	69.420%#	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10046	19.073	ug/l	96
20) Methylene Chloride	2.788	84	7107	6.050	ug/l	94
25) 2-Butanone	4.568	43	5410	7.156	ug/l #	89
37) Ethyl Acetate	4.715	43	15257	10.205	ug/l	100
43) Isopropyl Acetate	6.336	43	68432	28.447	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033478.D
Acq On : 19 Dec 2022 12:50
Operator : JC/MD
Sample : PB149705ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149705ZHE#02

Quant Time: Dec 20 05:00:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

