

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033479.D
 Acq On : 19 Dec 2022 13:13
 Operator : JC/MD
 Sample : PB149705ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149705ZHE#03

Quant Time: Dec 20 05:00:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174129	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	34088	48.909	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.820%
35) Dibromofluoromethane	5.385	113	37805	48.270	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	78570	47.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.920%
62) 4-Bromofluorobenzene	11.079	95	60699	48.453	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.900%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	9438	18.300	ug/l	99
20) Methylene Chloride	2.794	84	7158	6.223	ug/l #	86
25) 2-Butanone	4.562	43	5315	7.180	ug/l #	89
37) Ethyl Acetate	4.714	43	15631	10.829	ug/l	98
43) Isopropyl Acetate	6.336	43	67186	28.929	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033479.D
Acq On : 19 Dec 2022 13:13
Operator : JC/MD
Sample : PB149705ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149705ZHE#03

Quant Time: Dec 20 05:00:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

