

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033482.D
 Acq On : 19 Dec 2022 14:22
 Operator : JC/MD
 Sample : PB149722ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149722ZHE#11

Quant Time: Dec 20 05:00:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188153	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178518	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36746	49.713	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.420%
35) Dibromofluoromethane	5.385	113	40598	47.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	8.653	98	85255	48.157	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.079	95	66134	48.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%
Target Compounds						
16) Acetone	2.386	43	22645	41.402	ug/l	98
20) Methylene Chloride	2.794	84	5822	4.773	ug/l #	90
37) Ethyl Acetate	4.714	43	15778	10.116	ug/l	99
43) Isopropyl Acetate	6.336	43	73673	29.357	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033482.D
Acq On : 19 Dec 2022 14:22
Operator : JC/MD
Sample : PB149722ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149722ZHE#11

Quant Time: Dec 20 05:00:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

