

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033484.D
 Acq On : 19 Dec 2022 15:08
 Operator : JC/MD
 Sample : PB149722ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149722ZHE#13

Quant Time: Dec 20 05:01:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	179658	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36666	52.328	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.660%		
35) Dibromofluoromethane	5.391	113	40580	50.219	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.440%		
50) Toluene-d8	8.647	98	86027	50.891	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.780%		
62) 4-Bromofluorobenzene	11.079	95	66751	51.644	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.280%		
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	22718	43.816	ug/l	99
20) Methylene Chloride	2.794	84	5575	4.821	ug/l	89
25) 2-Butanone	4.586	43	3911	5.255	ug/l	93
37) Ethyl Acetate	4.720	43	16115	10.821	ug/l	99
43) Isopropyl Acetate	6.342	43	73713	30.762	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033484.D
Acq On : 19 Dec 2022 15:08
Operator : JC/MD
Sample : PB149722ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149722ZHE#13

Quant Time: Dec 20 05:01:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

