

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033486.D
 Acq On : 19 Dec 2022 15:54
 Operator : JC/MD
 Sample : PB149722ZHE#15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149722ZHE#15

Quant Time: Dec 20 05:01:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	107323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	177600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36385	52.846	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.700%		
35) Dibromofluoromethane	5.385	113	39889	49.936	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.880%		
50) Toluene-d8	8.647	98	81610	48.837	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.680%		
62) 4-Bromofluorobenzene	11.079	95	63044	49.341	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.680%		
Target Compounds						
						Qvalue
16) Acetone	2.379	43	28715	56.361	ug/l	100
18) Methyl Acetate	2.703	43	1484	1.006	ug/l #	87
20) Methylene Chloride	2.794	84	6490	5.712	ug/l	91
25) 2-Butanone	4.568	43	4516	6.175	ug/l	98
37) Ethyl Acetate	4.714	43	17230	11.704	ug/l	98
43) Isopropyl Acetate	6.336	43	73591	31.067	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033486.D
Acq On : 19 Dec 2022 15:54
Operator : JC/MD
Sample : PB149722ZHE#15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149722ZHE#15

Quant Time: Dec 20 05:01:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

