

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033488.D
 Acq On : 19 Dec 2022 16:40
 Operator : JC/MD
 Sample : PB149722ZHE#17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149722ZHE#17

Quant Time: Dec 20 05:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	186866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37236	51.579	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.160%		
35) Dibromofluoromethane	5.391	113	41193	49.011	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.020%		
50) Toluene-d8	8.647	98	87159	49.571	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.140%		
62) 4-Bromofluorobenzene	11.079	95	66793	49.683	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.360%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	28244	52.871	ug/l	99
20) Methylene Chloride	2.794	84	6711	5.633	ug/l	87
25) 2-Butanone	4.562	43	4221	5.505	ug/l	98
37) Ethyl Acetate	4.715	43	16424	10.603	ug/l	98
43) Isopropyl Acetate	6.342	43	74222	29.780	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033488.D
Acq On : 19 Dec 2022 16:40
Operator : JC/MD
Sample : PB149722ZHE#17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149722ZHE#17

Quant Time: Dec 20 05:01:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

