

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033492.D
 Acq On : 19 Dec 2022 18:36
 Operator : JC/MD
 Sample : N6097-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FVSUB-TP1-COMP-01

Quant Time: Dec 20 05:02:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108219	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178568	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37722	54.334	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.660%
35) Dibromofluoromethane	5.391	113	42815	53.308	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.620%
50) Toluene-d8	8.647	98	89198	53.088	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.180%
62) 4-Bromofluorobenzene	11.079	95	68075	52.990	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.980%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	15829	30.812	ug/l	97
20) Methylene Chloride	2.794	84	13404	11.699	ug/l #	85
37) Ethyl Acetate	4.715	43	14144	9.555	ug/l	99
43) Isopropyl Acetate	6.336	43	62995	26.450	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033492.D
Acq On : 19 Dec 2022 18:36
Operator : JC/MD
Sample : N6097-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FVSUB-TP1-COMP-01

Quant Time: Dec 20 05:02:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

