

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033493.D
 Acq On : 19 Dec 2022 18:59
 Operator : JC/MD
 Sample : N6098-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FVSUB-TP3-COMP-04

Quant Time: Dec 20 05:02:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190618	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36647	49.212	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.420%
35) Dibromofluoromethane	5.391	113	41395	48.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.560%
50) Toluene-d8	8.647	98	86558	48.261	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.520%
62) 4-Bromofluorobenzene	11.079	95	65881	48.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	31545	57.246	ug/l	99
20) Methylene Chloride	2.794	84	12062	9.815	ug/l	89
37) Ethyl Acetate	4.721	43	11267	7.131	ug/l	98
43) Isopropyl Acetate	6.342	43	54016	21.246	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033493.D
Acq On : 19 Dec 2022 18:59
Operator : JC/MD
Sample : N6098-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FVSUB-TP3-COMP-04

Quant Time: Dec 20 05:02:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

