

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
 Data File : VX033495.D
 Acq On : 19 Dec 2022 19:45
 Operator : JC/MD
 Sample : N6098-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FVSUB-TP2-COMP-02

Quant Time: Dec 20 05:02:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117536	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	36670	48.632	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.385	113	41597	47.488	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	8.647	98	85522	46.672	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.340%
62) 4-Bromofluorobenzene	11.079	95	66362	47.364	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	27023	48.431	ug/l	100
20) Methylene Chloride	2.788	84	14757	11.859	ug/l #	87
37) Ethyl Acetate	4.715	43	13376	8.286	ug/l	99
43) Isopropyl Acetate	6.336	43	61856	23.814	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121922\
Data File : VX033495.D
Acq On : 19 Dec 2022 19:45
Operator : JC/MD
Sample : N6098-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FVSUB-TP2-COMP-02

Quant Time: Dec 20 05:02:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

