

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121923\
 Data File : VX039370.D
 Acq On : 19 Dec 2023 20:30
 Operator : JC/MD
 Sample : VX1219MBL02
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK653

Quant Time: Dec 20 00:46:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 20 00:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	211693	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	196780	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	94516	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	86014	51.803	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.600%
7) Chloroethane-d5	1.672	69	62198	50.679	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.360%
11) 1,1-Dichloroethene-d2	2.306	65	40592	55.693	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	111.380%
21) 2-Butanone-d5	4.446	46	152496	101.461	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.460%
24) Chloroform-d	5.056	84	149896	50.997	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.952	65	108812	57.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.320%
32) Benzene-d6	5.970	84	322101	53.038	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.080%
36) 1,2-Dichloropropane-d6	7.306	67	102658	52.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.060%
41) Toluene-d8	8.647	98	285384	51.854	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.700%
43) trans-1,3-Dichloroprop...	8.945	79	46561	46.815	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.620%
47) 2-Hexanone-d5	9.384	63	109383	93.705	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.700%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	139115	50.176	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.360%
66) 1,2-Dichlorobenzene-d4	12.317	152	94093	52.668	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121923\
Data File : VX039370.D
Acq On : 19 Dec 2023 20:30
Operator : JC/MD
Sample : VX1219MBL02
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK653

Quant Time: Dec 20 00:46:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 20 00:14:00 2023
Response via : Initial Calibration

