

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121923\
 Data File : VX039381.D
 Acq On : 20 Dec 2023 00:40
 Operator : JC/MD
 Sample : 05869-07 100X
 Misc : 4.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AT3

Quant Time: Dec 20 02:22:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 20 00:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	176516	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	164413	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	78215	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	62457	45.112	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	90.220%		
7) Chloroethane-d5	1.672	69	47616	46.530	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	93.060%		
11) 1,1-Dichloroethene-d2	2.306	65	30028	49.409	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	98.820%		
21) 2-Butanone-d5	4.452	46	123596	98.621	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	98.620%		
24) Chloroform-d	5.056	84	118502	48.351	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.700%		
26) 1,2-Dichloroethane-d4	5.958	65	85168	53.657	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	107.320%		
32) Benzene-d6	5.970	84	245877	48.457	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.920%		
36) 1,2-Dichloropropane-d6	7.305	67	79981	48.979	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	97.960%		
41) Toluene-d8	8.647	98	214901	46.735	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.460%		
43) trans-1,3-Dichloroprop...	8.945	79	34418	41.418	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.840%		
47) 2-Hexanone-d5	9.384	63	87638	89.857	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	89.860%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	110453	47.681	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	95.360%		
66) 1,2-Dichlorobenzene-d4	12.317	152	72315	48.914	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.820%		
Target Compounds						Qvalue
51) Chlorobenzene	10.079	112	3771	1.102	ug/L	98
65) 1,4-Dichlorobenzene	12.036	146	1899	0.766	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121923\
Data File : VX039381.D
Acq On : 20 Dec 2023 00:40
Operator : JC/MD
Sample : 05869-07 100X
Misc : 4.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AT3

Quant Time: Dec 20 02:22:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 20 00:14:00 2023
Response via : Initial Calibration

