

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006668.D
 Acq On : 20 Dec 2018 11:47
 Operator : JC/MD
 Sample : J6474-01DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0492DL

Quant Time: Dec 21 03:57:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	651827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1025913	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	946978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	460839	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	353301	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
35) Dibromofluoromethane	5.50	113	306744	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
50) Toluene-d8	8.71	98	1216860	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.13	95	438622	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	61121	9.109	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	19940	3.090	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	296909	41.703	ug/l	91
44) Trichloroethene	7.21	130	475280	60.799	ug/l	95
64) Tetrachloroethene	9.33	164	5314	0.709	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006668.D
Acq On : 20 Dec 2018 11:47
Operator : JC/MD
Sample : J6474-01DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0492DL

Quant Time: Dec 21 03:57:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

