

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006675.D
 Acq On : 20 Dec 2018 14:54
 Operator : JC/MD
 Sample : J6484-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 969-MW-07(22.5)

Quant Time: Dec 21 04:12:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	624324	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	983077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	907225	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	434679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	342864	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	5.51	113	295230	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	8.71	98	1172428	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.13	95	426391	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
Target Compounds						
16) Acetone	2.44	43	11296	3.654	ug/l	90
64) Tetrachloroethene	9.34	164	21595	3.006	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006675.D
Acq On : 20 Dec 2018 14:54
Operator : JC/MD
Sample : J6484-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
969-MW-07(22.5)

Quant Time: Dec 21 04:12:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

