

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006677.D
 Acq On : 20 Dec 2018 15:47
 Operator : JC/MD
 Sample : J6484-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 971-MW-15(17)

Quant Time: Dec 21 04:15:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	620134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	979825	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	894608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	426454	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	343620	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
35) Dibromofluoromethane	5.50	113	287867	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
50) Toluene-d8	8.71	98	1160129	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	11.13	95	414460	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
Target Compounds						
64) Tetrachloroethene	9.34	164	2476	0.350	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006677.D
Acq On : 20 Dec 2018 15:47
Operator : JC/MD
Sample : J6484-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
971-MW-15(17)

Quant Time: Dec 21 04:15:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

