

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006679.D
 Acq On : 20 Dec 2018 16:41
 Operator : JC/MD
 Sample : J6484-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 974-MW-18(21.5)

Quant Time: Dec 21 04:18:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	613525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	978779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	900531	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	429394	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	334525	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
35) Dibromofluoromethane	5.50	113	290369	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
50) Toluene-d8	8.71	98	1168826	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	11.14	95	415622	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
Target Compounds						
16) Acetone	2.45	43	10980	3.614	ug/l #	86
27) cis-1,2-Dichloroethene	4.60	96	9906	1.478	ug/l	88
44) Trichloroethene	7.21	130	11181	1.499	ug/l	98
64) Tetrachloroethene	9.34	164	38361	5.380	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006679.D
Acq On : 20 Dec 2018 16:41
Operator : JC/MD
Sample : J6484-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
974-MW-18(21.5)

Quant Time: Dec 21 04:18:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

